

Myron L 300 QCP (NIST) conductivity calibration standard, 445 µS

Myron L – Mfr # 442-300Q-CP – N° d'article FV-01489-93

Conductivity values are accurate to within ±1% and are referenced to 25°C

- Standards have NaCl and 442™ conversions included for ease of use

Caractéristiques & Description

Conductivity Value (µS)	445
Resistivity Value (MΩ)	0.0022
Volume (mL)	946
Volume (oz)	32
Conductivity-to-TDS Calibration Values (ppm 442)	300

Conductivity-to-TDS Calibration Values (ppm NaCl)	214.3
Description	conductivity calibration standard, 445 µS