

Myron L 1500 QCP (NIST) conductivity calibration standard, 2060 µS

Myron L – Mfr # 442-1500Q-CP – N° d'article FV-01489-95

Conductivity values are accurate to within ±1% and are referenced to 25°C

- Standards have NaCl and 442™ conversions included for ease of use

Caractéristiques & Description

Conductivity Value (µS)	2060
Resistivity Value (MΩ)	0.00048
Volume (mL)	946
Volume (oz)	32
Conductivity-to-TDS Calibration Values (ppm 442)	1500

Conductivity-to-TDS Calibration Values (ppm NaCl)	1036
Description	conductivity calibration standard, 2060 µS