Cole-Parmer®
an antylia scientific company



Myron L 3K QCP (NIST) conductivity calibration standard, 3900 µS

Myron L – Mfr # 442-3000Q-CP – N° d'article FV-01489-96

Conductivity values are accurate to within ±1% and are referenced to 25°C

- Standards have NaCl and 442™ conversions included for ease of use

Caractéristiques & Description

Conductivity Value (µS)	3900
Resistivity Value (MΩ)	0.00025
Volume (mL)	946
Volume (oz)	32
Conductivity-to-TDS Calibration Values (ppm 442)	3000
Conductivity-to-TDS	2027

Calibration Values (ppm NaCl)	
Description	conductivity calibration standard, 3900 µS