

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

## Product Description

**Catalog Number:** 3324

**Product Description:** Corning® CellSTACK® - Universal Cap, 1 solid overcap, 2 vent caps

**Component Materials:**

- Caps - Dow LDPE 993I Natural
- Air foil - Selig 1-19/S70A,  
0.2µm GORE-TEX® PTFE membrane meets, *USP, Class VI* requirements for plastic containers and closures.

**Product Dimensions:**

- |                        |   |              |                        |   |           |
|------------------------|---|--------------|------------------------|---|-----------|
| Diameter of solid cap  | - | 0.690 in.    | Diameter of vent cap   | - | 0.770 in. |
| Height of solid cap    | - | 0.562 in.    | Height of vent cap     | - | 0.561 in. |
| Sm diameter of airfoil | - | 0.250 in.    | Lg diameter of airfoil | - | 0.750 in. |
| Tolerances             | - | +/- 0.05 in. | Spacing of airfoil     | - | 0.750 in. |

**Sterilization:**

The product is irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (AAMI TIR 27) *Sterilization of health care products-Requirements for validation and routine control-Radiation sterilization*.  
Sterility Assurance Level: SAL 10<sup>-6</sup>

**Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual examination of the product.

Packaging:

Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 2