

Catalog Number: 3329

Product Description: Corning® CellSTACK® Filling Accessory, Female MPC and plug

Component Materials:

Cap Virgin High Density Polyethylene, *meets USP, Class VI requirements* for plastic containers and closures. Heavy metal free (meets CONEG requirements) color concentrate.

Tubing Thermoplastic Elastomer, *meets USP, Class VI requirements* for plastic containers and closures

Female MPC and Plug Polycarbonate, *meets USP, Class VI requirements* for plastics.

Product Dimensions:

Inner Dimension - 0.375"

BSE/TSE - Product complies with the latest revision of EMA/410/01 "Note for Guidance on minimising the risk of transmitting animal spongiform encephalopathy agents via human and veterinary medicinal products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01.

Sterilization - Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, "Sterilization of health care products- Radiation". Products meet a minimum Sterility Assurance Level (SAL) of 10⁻⁶

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass
Packaging Inspection - Pass

Lot Number Designation:

8 Digit Lot Number: First 3 digits – Julian date, start of manufacturing; Next 2 digits – Year of manufacture; Last 3 digits – Batch identification.