

CORNING GOSSELIN SAS  
123 Route de Caestre  
59190 - BORRE  
FRANCE  
[www.corning.com/lifesciences](http://www.corning.com/lifesciences)  
Refer to website for regional contact information

Date of certificate : 2026 - 03 - 02

Page 1

### Product description

*Bottle Round HDPE 500mL, 58mm Red Cap Assembled, Sterile, 140/Case*

### Catalogue number

*LR500-02*

### Manufacturing date

*2026/02/01*

### Lot number

*968098*

### Quality Management System

Complies with the current version of the ISO9001 standard.

### Sterility Assurance Level

Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, "Sterilization of health care products - Radiation". Products meet a minimum Sterility Assurance Level (SAL) of 10<sup>-3</sup>.

### Expiration date

*2029/02/01*

### Expiration statement

*This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.*

### Performance testing (AQL)

*Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.*

This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release



Sandra Bianchi  
Plant Quality Manager, Borre  
[BianchiS@corning.com](mailto:BianchiS@corning.com)