

CORNING GOSSELIN SAS
123 Route de Caestre
59190 - Borre
FRANCE
www.corning.com/lifesciences
Refer to website for regional contact information

Date of certificate : 2025 - 11 - 24

Page 1

Product description

Bottle Square HDPE 2L 58mmRedCapSea / Assembled Sterile 50/CS

Catalogue number

LC2000-03

Lot number

960835

Manufacturing date

2025/10/01

Quality Management System

Complies with the current version of the ISO9001 standard.

Sterility Assurance Level

Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, "Sterilization of health care products - Radiation". Products meet a minimum Sterility Assurance Level (SAL) of 10⁻³.

Expiration date

2028/10/01

Expiration statement

This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release



Sandra Bianchi
Plant Quality Manager, Borre
BianchiS@corning.com