

Electrolyte Quality Control Certificate

Nominal Specifications

Electrolyte Solution	Electrolyte KCl 3 mol/l
Order Number(s)	51343180, 51350072, 51350080

Electrolyte Properties

Measured Potential Difference (mV)	0.5 (25°C)
Lot Number	1K075B
Production Date	15 March 2024
Expiry Date	15 March 2027
Electrolyte Substances	Potassium Chloride

Traceability

The measurement value on this certificate was determined by comparison to a secondary standard electrolyte. The secondary standard electrolyte is a recipe that was determined from measurements against NIST primary SRM 2202.

Calibration Procedure	One point calibration	
Unit of Measurement	mV	
Reference Material	Comparison Std. 3M KCl with AgCl (sat.)	
Reference Material Lot Number(s)	1J299G	
Reference Material Specification	0.9788	

©2019, Mettler-Toledo GmbH, Analytical, Im Langacher 44, 8606 Greifensee, Switzerland

METTLER TOLEDO

Date of Certificate Issue
15 March 2024



Quality Manager
Peter Rowing