Nolato Treff Certificate of Conformity

CoC Certificate according to DIN EN 10204 2.1

CleanRoom® PURE

Article:	Deep Well Plate PP 1.2 ml sterile
Article Number:	96.09118.9.01 / 034053
Purchase Number:	406863, 408793
Lot- Number:	5177456EXP2029-09-02
Quantity:	64 pcs

Statement

Nolato Treff AG confirms that TreffLab consumables and products marked with the CleanRoom® Pure quality symbol have been manufactured in microbiologically controlled clean rooms class 8, ISO 14644-1. To avoid direct human contact, employees are fully protective dressed and masked.

Quality Parameters

Quality standard for original sealed packages

- Free of RNase (< 5x10E-9 Kunitz units)
- Free of DNase (< 5x10E-6 Kunitz units)
- Free of DNA (< 2 pg DNA)
- Free of endotoxins (< 10 copies)
- Free of PCR inhibitors (< 0.001 EU/ml)
- Free of ATP (< 30 RLU)

Analyses are performed by an independent Swiss laboratory for molecular biological & microbiological analytics, accredited ISO/IEC17025 for real-time PCR.

Date: 4 August 2026
Place: CH-9113 Degersheim
Nolato Treff AG

Nolato

Nolato Treff AG | CH-9113 Degersheim

Customer Service
Julia Hungerbühler