

# Deepwell Plates and Microplates

## Eppendorf Deepwell Plates

Instructions for Use

Copyright © 2025 Eppendorf SE, Germany. All rights reserved, including graphics and images. No part of this publication may be reproduced without the prior permission of the copyright owner.

Eppendorf® and the Eppendorf Brand Design are registered trademarks of Eppendorf SE, Germany.

Eppendorf trademarks and trademarks of third parties may appear in this manual. All trademarks are the property of their respective owners. The respective trademark name, representations and listed owners can be found on [www.eppendorf.com/ip](http://www.eppendorf.com/ip).

U.S. Patents and U.S. Design Patents are listed on [www.eppendorf.com/ip](http://www.eppendorf.com/ip).

## **1 About this manual**

### **1.1 Notes on this manual**

When using this product in combination with other products or devices, observe the relevant operating manuals. This document does not replace the instructions for use provided with other products or devices.

1. Before using the product, read this document in full.
2. Make sure that this document is available to you while using the product.

The dates in this manual correspond to the international date format as specified in the ISO 8601 standard. All dates are shown in the format YYYY-MM-DD or YYYY-MM.

### **1.2 Other applicable documents**

The following documents supplement this manual:

- Technical Data Sheet for "Eppendorf Deepwell Plate 96/500  $\mu\text{L}$ "
- Technical Data Sheet for "Eppendorf Deepwell Plate 96/1000  $\mu\text{L}$ "
- Technical Data Sheet for "Eppendorf Deepwell Plate 96/2000  $\mu\text{L}$ "
- Technical Data Sheet for "Eppendorf Deepwell Plate 384/200  $\mu\text{L}$ "
- Instructions for use for "Eppendorf Storage Film, Storage Foil, sealing Mat and Plate Lid"

## **Safety**

### **4 Eppendorf Deepwell Plates English (EN)**

## **2 Safety**

### **2.1 Intended use**

Eppendorf Deepwell Plates and Microplates are uncoated polypropylene single-use items for the analysis, preparation, mixing, centrifugation, transport, and storage of solid and liquid samples and reagents. Eppendorf Deepwell Plates and Microplates are suitable for use in training, routine, and research laboratories in the chemical, industrial or Life Sciences sectors.

The product may only be used for research purposes. Eppendorf accepts no liability for the use of the product in other applications. It is not intended for use in diagnostic or therapeutic applications. Use of the Eppendorf Deepwell Plates and Microplates requires personnel who are trained and skilled in the areas mentioned above. Eppendorf Microplates are suitable for use in all standard plate readers.

### **2.2 Personal protective equipment**

Personal protective equipment serves to ensure the safety and protection of the user when working with the product.

Personal protective equipment must comply with country-specific regulations and the regulations of the laboratory.

### **2.3 Residual risks when used as intended**

If the product is not used as intended, the installed safety devices may not function correctly. To reduce the risk of personal injury and material damage and to avoid dangerous situations, please observe the general safety instructions.

#### **2.3.1 Personal injury**

##### **2.3.1.1 Biological hazards**

Poisonous, radioactive or aggressive chemicals, as well as infectious liquids and pathogenic biological agents, can harm your health and damage the environment.

- Observe the national regulations and the biosafety level of your laboratory when handling these substances.
- Wear your personal protective equipment.
- Observe the Safety Data Sheets and instructions for use from the manufacturer.
- Read the "Laboratory Biosafety Manual" (source: World Health Organization, Laboratory Biosafety Manual, in its current version) about handling germs or biological material of risk group II or higher.

Risk of contamination when disposing of used consumables.

- Dispose of used consumables in accordance with the substances with which they have come into contact.
- Observe the rules that apply to your laboratory.

### 2.3.1.2 Chemical hazards

Leaking substances may present a risk to health.

- Only use the product in a visually perfect and undamaged condition from intact packing.
- Observe the maximum filling volume.
- Remove seals carefully. When removing the selected sealing system, liquid may squirt out.
- Seal the plates before centrifugation. Observe the information in the operating manual provided with the device.
- Store and transport the product in such a way that ensures substances do not leak from the product.
- Do not use liquid nitrogen in combination with this product. The product and the seals may become damaged during the thawing process and burst.
- Do not use the product as a cryogenic tube.

Samples may become contaminated if the plates are used multiple times.

- Only use the plates once.
- Dispose of the plates after using them just once. Comply with the corresponding disposal regulations for the substances and samples used.
- Pay attention to the rules applicable for your laboratory.

Risk of contamination from damaged plates during centrifugation.

During centrifugation, the plates are exposed to heavy loads. If used incorrectly, they may become damaged and release their substances.

- Observe the maximum permitted centrifugation forces.
- Close the plates carefully.
- Read the operating manual provided for the centrifuge used.
- Centrifuge stacked plates at a low speed only.
- Please note that organic solvents may reduce the mechanical strength of the plates. If in doubt, contact your contact person.
- Carry out a test run to determine maximum centrifugation stability. Reduce the centrifugation forces for the test run.

### **2.3.2 Material damage**

#### **2.3.2.1 Incorrect handling**

##### **Sample contamination from open or damaged packing**

The designated purity of the consumables is only guaranteed with sealed packing.

- Check that the packing is undamaged.
- Only open the packing immediately before use.
- Only use visually perfect items.

##### **Use in incorrect temperature range**

Extreme temperatures (e.g. when deep-freezing or autoclaving) influence the material. Mechanical strength, dimensions and the shape of the consumable change.

- Only use consumables that are suitable for the temperature range and procedure selected.

##### **Sample loss due to incorrect equipment**

Please note that the use of plates in incorrect equipment (e.g. incorrect rotor/adaptor or mixer insert) may destroy the plates and lead to the plate contents leaking.

- Take note of the height and diameter of the plates. Only use suitable centrifuge inserts or mixer thermoblocks.
- Use the adapters intended for these plates.

### **3 Product description**

Eppendorf Deepwell Plates and Microplates are uncoated polypropylene single-use items for the analysis, preparation, mixing, centrifugation, transport, and storage of solid and liquid samples and reagents. Eppendorf Microplates are suitable for use in all standard plate readers. The use of the plates for UV/VIS applications is not recommended. Eppendorf Deepwell Plates and Microplates are characterized by a high chemical and temperature resistance and can be sealed when suitable sealing options are used. The plates meet the Microplate standards ANSI/SLAS 1-2004 to ANSI/ SLAS 4-2004.

#### **3.1 Plate variants**

##### **BioBased Consumables**

BioBased Consumables are single-use vessels made of bio-based raw materials. The polypropylene comes from renewable sources (e.g. waste or vegetable oil residues) that are reprocessed and put to new use.

These products have the same quality and performance features as Eppendorf single-use vessels made of fossil raw materials. The processed quantities of bio-based raw materials are declared and subjected to inspection by an independent system. As a participant in this system, Eppendorf is certified.

You can find information on the bio-based raw material content according to the mass balance approach on the label of the direct packing. For further information, refer to the respective product pages at [www.eppendorf.com](http://www.eppendorf.com).

The BioBased Consumables are non-biodegradable and non-compostable. Please observe the official disposal regulations in your laboratory.

##### **LoBind**

The LoBind Plates have been developed to facilitate a better recovery rate. The material used for this purpose allows for a lower loss of target molecules. The performance of the LoBind effect may vary considerably depending on the specific application.

##### **Forensic DNA Grade**

The Forensic DNA Grade Plates have been designed to meet the strict requirements of forensic applications. They comply with ISO 18385. Testing is carried out by an external test laboratory (accredited according to ISO 17025). Lot-specific certificates record the purity criteria.

#### **3.2 Barcoded plates**

The Eppendorf twin.tec PCR Plates are also available with barcodes. Eppendorf SafeCode Plates are provided with barcodes and are available from stock. There is also the option of applying a customer-specific barcode.

**8 Product description**  
 Eppendorf Deepwell Plates  
 English (EN)

<b>Pre-barcoded plates (SafeCode)</b>	<b>Customer-specific barcoded plates</b>
Pre-barcoded	Customer-specific barcodes
Available from stock	Production on request
Print	Label
ep-unique ID (serial number); unique for all Eppendorf SafeCode Consumable	Customer-specific ID (serial number)
Code 128 (2 letters + 10 digits)	Customer-specific choice of code 128 (8 or 10 digits), code 39, or Interleaved 2 of 5 Optional prefix
2D DataMatrix Code on front side 1D barcode on both short sides and front side Plain text on front side	1D barcode and plain text as specified by customer possible on all four sides.
SafeCode Feature	No SafeCode Feature

### 3.3 SafeCode plates

The SafeCode plates come with a code in the form of an ID. This ID corresponds to a serial number and is unique across all Eppendorf SafeCode Consumables, lots, Vials, tubes and plates. This facilitates a precise tracking of samples in the laboratory.

This ID links all the relevant production data and product-specific documents, such as lot number, technical data and certificate of conformity. This data can be accessed via the Eppendorf webpage at [www.eppendorf.com/safecode-data](http://www.eppendorf.com/safecode-data) (Service & Support > Quality & Certificates > Certificates).

#### Codes

The serial number is coded on the front with a 2D DataMatrix code and a 1D code. The 1D code is also located on the two short sides. The code is printed in black on a white background to ensure legibility and suitability for use in machinery, and is particularly resistant to scratching.

Eppendorf uses the Reed Solomon algorithm (ECC200 - Error Correcting Code) for the error correction of the DataMatrix code. The ECC200 DataMatrix symbology ensures that DataMatrix codes are still legible with surface damage of up to 25%. This code is unique within the Eppendorf SafeCode-Consumables portfolio and allows for the clear identification of samples. Use of the SafeCode-Consumables is also facilitated by the plain text (human readable code).

#### SafeCodes

The ep-unique ID SafeCode is a serial number issued by Eppendorf that is unique to every Eppendorf SafeCode Consumable. This also prevents any mix-up across lots or products (Eppendorf Vials, vessels and plates).

Resolution	14×14 pixels
Marking	QR code + ep-unique ID
ep-unique-ID format	ep + 10 digits
Availability	From stock
Code type	2D: Data Matrix, ECC200 1D: Code 128
Uniqueness	Unique for every Eppendorf SafeCode Consumable
ISO standard	Codes: ISO/IEC 16022, ISO/IEC 15417 Printed material: ISO/IEC 15416, ISO/IEC 29158

**Product description**

Eppendorf Deepwell Plates  
English (EN)



SafeCodes may become damaged and illegible due to scratching, fading, autoclaving and other influences.

SafeCode Consumables are not suitable for autoclaving. Autoclaving damages the SafeCode and can result in a loss of information.

**Use**

Check the compatibility of your system (reading device and software) before using the various SafeCode Consumables. Also check whether the data transfer in your systems is working correctly.



Multiple use of SafeCode Consumables can generate inconsistent data. This can lead to a loss of information and a mix-up of samples. Consider the consequences of losing information or mixing up the samples. Implement measures to reduce this risk where appropriate.

**3.4 Sealing**

The sealing quality depends on the interaction of the individual system components, including the geometry and polymer material of the plates, the material of the sealing film or foil, the sealing method used and, in the case of heat sealing, the sealing time and temperature.

When applying adhesive seals manually, we recommend using a squeegee or similar tool to ensure adequate pressure during the sealing process.

For automatic sealing, please refer to the instructions for use of your sealing device.

Recommended sealing parameters for the use of Eppendorf HeatSealers and Eppendorf Heat Sealing Films and Foils can be found in the document "HeatSealer S100 and S200 Sealing recommendations".



Multiple sealing of the plate can lead to reduced seal tightness and increased sample evaporation. If applicable, check the tightness of the seal for multiple sealing in the specific customer protocol.

**4 Technical data**

	<b>Deepwell plates</b>			
	<b>96/2000 <math>\mu</math>L</b>	<b>96/1000 <math>\mu</math>L</b>	<b>96/500 <math>\mu</math>L</b>	<b>384/200 <math>\mu</math>L</b>
Material	Polypropylene (colored, colorless)			
Chemical resistance	See Application No. 56: The best material for original Eppendorf Tubes® and Plates and User Guide Nr. 23: Chemical Stability of Consumables in our download center at <a href="http://www.eppendorf.com/manuals">www.eppendorf.com/manuals</a> .			
Dimensions	As per ANSI/SLAS 1-2004 bis ANSI/SLAS 4-2004 (SLAS: Society for Laboratory Automation and Screening)			
Bottom shape	Conical	Round	Round	Conical
Total theoretical volume per well	2400 $\mu$ L	1200 $\mu$ L	700 $\mu$ L	240 $\mu$ L
Working volume per well	50 $\mu$ L – 2000 $\mu$ L	30 $\mu$ L – 1000 $\mu$ L	30 $\mu$ L – 550 $\mu$ L	20 $\mu$ L – 225 $\mu$ L
Autoclavability	121 °C, 20 min, unsealed.  The stability of the single-use items may be compromised. Autoclaving is not recommended for the SafeCode Consumables and customer-specific barcoded plates, as the durability of the marking may be impaired.			
Storage before use	Protect against sunlight and UV light. Store in a cool, dry place.			
Sample storage	Store plates containing samples in an upright position. The maximum filling volume at low temperatures must not exceed 80% of the nominal volume.			
	1600 $\mu$ L	800 $\mu$ L	440 $\mu$ L	180 $\mu$ L
Operating temperature	-86 °C – 100 °C			

**Technical data**

Eppendorf Deepwell Plates

English (EN)

	Microplates				
	96/F	96/U	96/V	384/F	384/V
Material	Polypropylene (colored, colorless)				
Chemical resistance	See Application No. 56: The best material for original Eppendorf Tubes® and Plates and User Guide Nr. 23: Chemical Stability of Consumables in our download center at <a href="http://www.eppendorf.com/manuals">www.eppendorf.com/manuals</a> .				
Dimensions	As per ANSI/SLAS 1-2004 bis ANSI/SLAS 4-2004 (SLAS: Society for Laboratory Automation and Screening)				
Bottom shape	Flat	Round	Conical	Flat	Conical
Total theoretical volume per well	400 µL	360 µL	350 µL	150 µL	140 µL
Working volume per well	50 µL – 350 µL	2 µL – 320 µL	20 µL – 300 µL	10 µL – 120 µL	5 µL – 120 µL
Autoclavability	121 °C, 20 min, in unsealed state. The stability of the single-use items may be compromised. Autoclaving is not recommended for the SafeCode Consumables and customer-specific barcoded plates, as the durability of the marking may be impaired.				
Storage before use	Protect against sunlight and UV light. Store in a cool, dry place.				
Sample storage	Store plates containing samples in an upright position. The maximum filling volume at low temperatures must not exceed 80% of the nominal volume.				
	280 µL	256 µL	240 µL	96 µL	96 µL
Operating temperature	-86 °C – 100 °C				

Current certificates can be found at [www.eppendorf.com](http://www.eppendorf.com).

**4.1 Centrifugation stability**

The centrifugation stability of consumables generally depends on the following conditions:

- Properties of the consumable (e.g., material, shape)
- Combination of centrifuge, rotor and, if applicable, adapter
- Fitting accuracy of the consumable in the rotor bore or adapter

- Centrifugation parameters (rotational speed/g-force, temperature, centrifugation time)
- Overall weight of consumable and contents
- Physical and chemical properties of the centrifuged liquid

Check the integrity of the consumable after centrifugation.



In non-refrigerated centrifuges, the temperature in the rotor chamber, rotor, and sample can rise above 40 °C, depending on the run time, g-force (rcf) / speed, and ambient temperature.

- Please note that this will reduce the centrifugation stability of the micro test tubes and plates.
- Take note of the temperature resistance of the samples.



The mechanical strength of the tubes is reduced by the use of organic solvents.

- In order to determine the maximum centrifugation stability for your applications, carry out a test run with lower g-force.

#### Centrifugability at 4 °C – 40 °C

The plates can be centrifuged under the following conditions at the maximum g-forces (rcf) listed in the table below:

- in a swing-bucket rotor
- 40 °C sample temperature with diluted saline solution (density 1.0 g/mL)
- 90 min

Purity grade	Eppendorf Deepwell Plates	Eppendorf Microplates
Sterile	5000 × g	6000 × g
PCR clean	6000 × g	6000 × g

Product versions	Eppendorf Deepwell Plates	Eppendorf Microplates
BioBased	6000 × g	N. A.
DNA LoBind	6000 × g	6000 × g
Protein LoBind	6000 × g	6000 × g




**14** **Technical data**  
**Eppendorf Deepwell Plates**  
**English (EN)**

The centrifugation stability of every plate type generally depends on the plate itself, on the centrifuge, rotor, rotor suspension, adapter, maximum revolution/maximum gravitational field, operating temperature, centrifugation time, density of centrifuged liquid, organic solvents, total cumulative weight, and correct operation.




**4.2 Purity grades**

**Deepwell plates and microplates**

The consumables are available in the following purity grades and purity criteria.

	PCR clean	Sterile	PCR clean + Protein-free
			
Eppendorf Deepwell Plates	■	■	
Eppendorf Deepwell Plates BioBased			■
Eppendorf Microplates	■	■	
DNA LoBind Plates	■		
Protein LoBind Plates	■		

Certified batch testing is carried out for the following purity grades and purity criteria.

	PCR clean	Sterile	PCR clean + Protein-free
			
Human DNA-free	■		■
DNase-free	■		■
RNase-free	■		■
PCR inhibitor-free	■		■
Pyrogen-free (endotoxin-free)		■	
Sterile (Ph.Eur./USP)		■	
Protein-free			■

In addition to internal process controls, each batch of products with the Steril, Protein-free, PCR clean or Biopur purity grades is tested and certified by an accredited external laboratory. Batch-related certificates are available online at: [www.eppendorf.com/certificates](http://www.eppendorf.com/certificates).

**Ordering information for deepwell plates**

Eppendorf Deepwell Plates  
English (EN)

**5 Ordering information for deepwell plates**

Eppendorf Deepwell Plates and Microplates are also available as the following options:

- Deepwell Plates: other border colors (yellow, green or blue)
- Bulk packages
- Available with barcodes on request (<https://www.eppendorf.com/barcodewizard>).

You can find ordering information on suitable sealing options at [www.eppendorf.com](http://www.eppendorf.com). Further product information is available in our product catalog, on our website [www.eppendorf.com](http://www.eppendorf.com) and at Eppendorf Application Support <https://www.eppendorf.com/application-support>.

**96/500  $\mu$ L**

Description	Order no.
<b>Deepwell Plate 96/500 <math>\mu</math>L</b>	
wells clear, 500 $\mu$ L	
PCR clean, white, 40 plates (5 bags $\times$ 8 plates)	0030 501 101
PCR clean, yellow, 40 plates (5 bags $\times$ 8 plates)	0030 501 110
PCR clean, green, 40 plates (5 bags $\times$ 8 plates)	0030 501 136
PCR clean, blue, 40 plates (5 bags $\times$ 8 plates)	0030 501 144
sterile, white, 40 plates (5 bags $\times$ 8 plates)	0030 502 108
sterile, yellow, 40 plates (5 bags $\times$ 8 plates)	0030 502 116
sterile, green, 40 plates (5 bags $\times$ 8 plates)	0030 502 132
sterile, blue, 40 plates (5 bags $\times$ 8 plates)	0030 502 140
PCR clean, white, 120 plates (10 bags $\times$ 12 plates)	0030 505 107
sterile, white, 120 plates (10 bags $\times$ 12 plates)	0030 506 103

**96/1000 µL**

Description	Order no.
<b>Deepwell Plate 96/1000 µL</b>	
wells clear, 1,000 µL	
PCR clean, white, 20 plates (5 bags × 4 plates)	0030 501 209
PCR clean, yellow, 20 plates (5 bags × 4 plates)	0030 501 217
PCR clean, green, 20 plates (5 bags × 4 plates)	0030 501 233
PCR clean, blue, 20 plates (5 bags × 4 plates)	0030 501 241
sterile, white, 20 plates (5 bags × 4 plates)	0030 502 205
sterile, yellow, 20 plates (5 bags × 4 plates)	0030 502 213
sterile, green, 20 plates (5 bags × 4 plates)	0030 502 230
sterile, blue, 20 plates (5 bags × 4 plates)	0030 502 248
PCR clean, white, 80 plates (10 bags × 8 plates)	0030 505 204
sterile, white, 80 plates (10 bags × 8 plates)	0030 506 200

**96/2000 µL**

Description	Order no.
<b>Deepwell Plate 96/2000 µL</b>	
wells clear, 2,000 µL	
PCR clean, white, 20 plates (5 bags × 4 plates)	0030 501 306
PCR clean, yellow, 20 plates (5 bags × 4 plates)	0030 501 314
PCR clean, green, 20 plates (5 bags × 4 plates)	0030 501 330
PCR clean, blue, 20 plates (5 bags × 4 plates)	0030 501 349
sterile, white, 20 plates (5 bags × 4 plates)	0030 502 302
sterile, yellow, 20 plates (5 bags × 4 plates)	0030 502 310
sterile, green, 20 plates (5 bags × 4 plates)	0030 502 337
sterile, blue, 20 plates (5 bags × 4 plates)	0030 502 345
PCR clean, white, 80 plates (10 bags × 8 plates)	0030 505 301
sterile, white, 80 plates (10 bags × 8 plates)	0030 506 308

**Ordering information for deepwell plates**

Eppendorf Deepwell Plates

English (EN)

**384/200 µL**

<b>Description</b>	<b>Order no.</b>
<b>Deepwell Plate 384/200 µL</b>	
wells clear, 200 µL	
PCR clean, white, 40 plates (5 bags × 8 plates)	0030 521 102
sterile, white, 40 plates (5 bags × 8 plates)	0030 522 109
PCR clean, white, 120 plates (10 bags × 12 plates)	0030 525 108
sterile, white, 120 plates (10 bags × 12 plates)	0030 526 104

**DNA LoBind 96/500 µL**

<b>Description</b>	<b>Order no.</b>
<b>Deepwell Plate 96/500µL</b>	
DNA LoBind®, wells clear, 500 µL, LoBind®	
PCR clean, white, 40 plates (5 bags × 8 plates)	0030 503 104
PCR clean, blue, 40 plates (5 bags × 8 plates)	0030 503 147
PCR clean, white, 120 plates (10 bags × 12 plates)	0030 507 100

**DNA LoBind 96/1000 µL**

<b>Description</b>	<b>Order no.</b>
<b>Deepwell Plate 96/1000µL</b>	
DNA LoBind®, wells clear, 1,000 µL, LoBind®	
PCR clean, white, 20 plates (5 bags × 4 plates)	0030 503 201
PCR clean, blue, 20 plates (5 bags × 4 plates)	0030 503 244
PCR clean, white, 80 plates (10 bags × 8 plates)	0030 507 207

**DNA LoBind 384/200 µL**

Description	Order no.
<b>Deepwell Plate 384/200µL</b> DNA LoBind®, wells clear, 200 µL, LoBind® PCR clean, white, 40 plates (5 bags × 8 plates)	0030 523 105
PCR clean, white, 120 plates (10 bags × 12 plates)	0030 527 100

**Protein LoBind 96/500 µL**

Description	Order no.
<b>Deepwell Plate 96/500 µL</b> Protein LoBind®, wells colorless, 500 µL PCR clean, white, 40 plates (5 bags × 8 plates)	0030 504 100
PCR clean, yellow, 40 plates (5 bags × 8 plates)	0030 504 119
PCR clean, white, 120 plates (10 bags × 12 plates)	0030 508 106

**Protein LoBind 96/1000 µL**

Description	Order no.
<b>Deepwell Plate 96/1000 µL</b> Protein LoBind®, wells colorless, 1,000 µL PCR clean, white, 20 plates (5 bags × 4 plates)	0030 504 208
PCR clean, yellow, 20 plates (5 bags × 4 plates)	0030 504 216
PCR clean, white, 80 plates (10 bags × 8 plates)	0030 508 203

**Protein LoBind 96/2000 µL**

Description	Order no.
<b>Deepwell Plate 96/2000 µL</b> Protein LoBind®, wells colorless, 2,000 µL PCR clean, white, 20 plates (5 bags × 4 plates)	0030 504 305

**Ordering information for deepwell plates**

Eppendorf Deepwell Plates  
English (EN)

**Protein LoBind 384/200 µL**

Description	Order no.
<b>Deepwell Plate 384/200µL</b> Protein LoBind®, wells colorless, 200 µL PCR clean, white, 40 plates (5 bags × 8 plates)	0030 524 101
PCR clean, white, 120 plates (10 bags × 12 plates)	0030 528 107

**Deepwell plates with SafeCode**

Description	Order no.
<b>Deepwell Plate 384/200 µL</b> wells clear, 200 µL, 2D SafeCode PCR clean, white, 40 plates (5 bags × 8 plates)	0030 113 551
<b>Deepwell Plate 96/1000 µL</b> wells clear, 1,000 µL, 2D SafeCode PCR clean, white, 20 plates (5 bags × 4 plates)	0030 113 535
<b>Deepwell Plate 96/2000 µL</b> wells clear, 2,000 µL, 2D SafeCode PCR clean, white, 20 plates (5 bags × 4 plates)	0030 113 527
<b>Deepwell Plate 96/500 µL</b> wells clear, 500 µL, 2D SafeCode PCR clean, white, 40 plates (5 bags × 8 plates)	0030 113 543

**Deepwell plates, biobased**

Description	Order no.
<b>Eppendorf Deepwell Plates BioBased 96/2000 µL</b> wells clear, 2,000 µL PCR clean and protein-free, white, 20 plates (5 bags × 4 plates)	0030 501 012
PCR clean and protein-free, spring green, 20 plates (5 bags × 4 plates)	0030 501 020
PCR clean and protein-free, crystal blue, 20 plates (5 bags × 4 plates)	0030 501 039

6 Ordering information for microplates

Eppendorf Assay/Reader Microplates

Description	Order no.
<b>Microplate 384/V</b> wells white PCR clean, gray, 80 plates (5 bags × 16 plates)	0030 621 670
wells black PCR clean, white, 80 plates (5 bags × 16 plates)	0030 621 905
<b>Microplate 96/F</b> wells white PCR clean, gray, 80 plates (5 bags × 16 plates)	0030 601 475
wells black PCR clean, white, 80 plates (5 bags × 16 plates)	0030 601 700
<b>Microplate 96/U</b> wells white PCR clean, gray, 80 plates (5 bags × 16 plates)	0030 601 572
wells black PCR clean, white, 80 plates (5 bags × 16 plates)	0030 601 807
<b>Microplate 96/V</b> wells white PCR clean, gray, 80 plates (5 bags × 16 plates)	0030 601 670
wells black PCR clean, white, 80 plates (5 bags × 16 plates)	0030 601 904

**Ordering information for microplates**

Eppendorf Deepwell Plates  
English (EN)

**Eppendorf Microplates**

<b>Description</b>	<b>Order no.</b>
<b>Microplate 384/F</b> wells clear, RecoverMax® well design PCR clean, white, 80 plates (5 bags × 16 plates) sterile, white, 80 plates (5 bags × 16 plates)	 0030 621 107 0030 622 103
<b>Microplate 384/V</b> wells clear, RecoverMax® well design PCR clean, white, 80 plates (5 bags × 16 plates) sterile, white, 80 plates (5 bags × 16 plates)	 0030 621 301 0030 622 308
<b>Microplate 96/F</b> wells clear, RecoverMax® well design PCR clean, white, 80 plates (5 bags × 16 plates) sterile, white, 80 plates (5 bags × 16 plates)	 0030 601 106 0030 602 102
<b>Microplate 96/U</b> wells clear, RecoverMax® well design PCR clean, white, 80 plates (5 bags × 16 plates) sterile, white, 80 plates (5 bags × 16 plates)	 0030 601 203 0030 602 200
<b>Microplate 96/V</b> wells clear, RecoverMax® well design PCR clean, white, 80 plates (5 bags × 16 plates) sterile, white, 80 plates (5 bags × 16 plates)	 0030 601 300 0030 602 307

### DNA LoBind

Description	Order no.
<b>Microplate 384/V-PP</b> DNA LoBind®, wells clear, 140 µL, LoBind® PCR clean, white, 80 plates (5 bags × 16 plates)	0030 623 304
PCR clean, white, 240 plates (10 bags × 24 plates)	0030 627 300
<b>Microplate 96/V-PP</b> DNA LoBind®, wells clear, 350 µL, LoBind® PCR clean, white, 80 plates (5 bags × 16 plates)	0030 603 303

### Protein LoBind

Description	Order no.
<b>Microplate 384/V-PP</b> Protein LoBind®, wells clear, 140 µL PCR clean, white, 80 plates (5 bags × 16 plates)	0030 624 300
PCR clean, white, 240 plates (10 bags × 24 plates)	0030 628 306

### Microplates with SafeCode

Description	Order no.
<b>Microplate 384/V 140 µL</b> wells clear, 140 µL, 2D SafeCode, RecoverMax® well design PCR clean, white, 80 (5 × 16)	0030 113 594
<b>Microplate 96/V</b> wells clear, 350 µL, 2D SafeCode, RecoverMax® well design PCR clean, white, 80 plates (5 bags × 16 plates)	0030 113 586

DOMINIQUE DUTSCHER

DOMINIQUE DUTSCHER

# Evaluate Your Manual

Give us your feedback.

[www.eppendorf.com/manualfeedback](http://www.eppendorf.com/manualfeedback)

DOMINIQUE DUTSCHER

**Your local distributor: [www.eppendorf.com/contact](http://www.eppendorf.com/contact)**

Eppendorf SE · Barkhausenweg 1 · 22339 Hamburg · Germany  
[eppendorf@eppendorf.com](mailto:eppendorf@eppendorf.com) · [www.eppendorf.com](http://www.eppendorf.com)