Technical Data Sheet

Edition Date: 2004-02-05
Revision Date : 2017-09-20
Revised by: Annette Roy, QA Coordinator

Product Name: Micrewtube®	**Catalogue No.: T332-4**
This document replaces any previous version	

1. Product Description:

- **1.5ml Micrewtube®:** Non sterile, self-standing tube with Silicone washer seal screw cap and attachment loop. Cap loops are pre-attached but caps are not screwed on. Tubes are non-graduated.

2. Packaging:

- **Case:** 1 000 units per case

3. Product Specifications:

- Material:
  - Polypropylene tube
  - Polypropylene cap
  - Silicone washer seal
- Certified RNase, DNase, Pyrogen and DNA Free
- Temperature range: -196°C to +121°C.
- Autoclavable at 121°C, for up to 30 minutes
- Centrifuged up to 17 000g

4. Standards and Conformity:

- **ISO 2859-1:** Sampling and inspection procedures
- **FDA:** Resin conforms to FDA 21 CFR 177.1520
- **USP:** Resin conforms to USP Class VI
- **CONEG:** Plastics and colorants are in conformity with CONEG standards for heavy metals
- **REACH (SVHC):** Plastic is in conformity to REACH standards
- **LATEX:** Material is Latex Free
- **BSE / TSE:** Material is BSE / TSE Free

5. Quality Assurance:

- Clear, no presence of contamination in plastic
- Visual attributes
- Volume measurements
- Closure verification
- Leak proof testing in vacuum at 71.3cm Hg
- Gas phase of Liquid Nitrogen resistance

DOMINIQUE DUTSCHER SAS

6. Traceability:

- **Lot No. Composition:** 8 or 9 digits

- **The lot number can be found in one or all of these locations:**

1. On exterior case label
2. On label inserted inside the master case
3. On the inner bag

7. Storage Conditions:

- Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

8. Recommended Use:

- Verify proper cap closure when using biohazard material and / or chemical reagents
- Follow chemical resistance chart recommendations
- For use in automated equipment, follow the equipment manufacturer's instructions
- Should be used only in the gas phase of Liquid Nitrogen

***OWNERSHIP OF MATERIALS**:*

*Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.*

DOMINIQUE DUTSCHER SAS