



PureProteome™ Albumin/IgG Depletion Kit

User Guide

Introduction

Human serum or plasma is a rich source of proteomic information, and is often interrogated for protein biomarkers of physiological and disease states. One of the overwhelming challenges in analyzing human serum (HS) is the wide concentration range of proteins present. Abundantly expressed proteins such as albumin and Immunoglobulin G (IgG) make up approximately 75% of the total protein in serum/plasma, while protein biomarkers may be present at much lower concentrations (ng/mL to pg/mL). High abundant proteins are a challenge for analytical methods, such as two-dimensional gel electrophoresis and mass spectrometry, because they mask the lower abundant proteins of interest. It is critical for these applications that high abundant proteins are efficiently, reproducibly, and specifically removed from serum samples, enabling accurate analysis of the lower abundant proteins.

The PureProteome Albumin/IgG Magnetic Beads have been developed using an antibody ligand specific for human serum albumin, and Protein G for the capture of IgG. These magnetic beads provide a rapid, scalable, and reproducible means to deplete > 98% of both albumin and IgG from serum and plasma samples, facilitating the detection and analysis of proteins of interest. PureProteome Magnetic Beads in combination with the PureProteome Magnetic Stand readily facilitate the depletion of multiple samples in parallel.

Kit Components

PureProteome Albumin/IgG Depletion Kit **Cat. No. LSKMAGD12**

PureProteome Albumin/IgG Magnetic Beads, 12 mL

Amicon® Ultra-4 3K Centrifugal Filter Device, 8 pk

10X Phosphate Buffered Saline (PBS), 7 mL

Materials Required to Use PureProteome Albumin/IgG Magnetic Beads

- For optimal performance, the Millipore PureProteome Magnetic Stand is recommended for use with PureProteome Magnetic Beads.
- 2 mL microcentrifuge tubes are required, but not provided.

LSKMAGD12MAN, Rev. A, 05/10

© 2010. Millipore Corporation. All rights reserved.

Millipore, Amicon, Milli-Q, and ZipTip are registered trademarks of Millipore Corporation. The M mark and PureProteome are trademarks of Millipore Corporation.

Application Guidelines

Please read the User Guide completely before beginning the protocol.

Albumin and IgG Depletion from Serum Samples

This protocol is optimized for 25 µL of serum. It may be scaled up or down as required by available sample volumes.

1. Prepare a 1X PBS working solution from the 10X PBS provided, by adding 1 mL of 10X PBS to 9 mL of Milli-Q® water. The working solution of 1X PBS will be used as the binding and wash buffer in the depletion protocol.
2. Mix the bead suspension so that all of the beads are uniformly resuspended. To ensure consistent bead volume, continue to mix while pipetting.
3. Pipette 950 µL of the resuspended beads into a 2 mL microcentrifuge tube. Place the tube into the PureProteome Magnetic Stand and allow the beads to migrate to the magnet. Remove the storage buffer with a pipette and discard.
4. Wash the beads twice, using 500 µL of 1X PBS for each wash. Disengage the magnet from the stand and vortex vigorously for 10 seconds. Re-engage the magnet and allow the beads to migrate to the magnet. Remove the buffer with a pipette and discard.
5. Dilute 25 µL of serum to a final volume of 100 µL with 1X PBS.
6. Add the diluted serum sample to the beads. Incubate for 60 minutes at room temperature with continuous mixing or end-over-end rotation.
7. Place each tube back into the magnetic stand. Allow the beads to migrate to the magnet. Remove the depleted serum with a pipette. Transfer to a fresh tube and save.
Note: This represents the albumin and IgG-depleted serum sample. The majority of the unbound proteins will be in the depleted serum/plasma sample.
8. To maximize recovery of the depleted serum sample, wash the beads 3 times, using 500 µL of 1X PBS for each wash. After each wash, disengage the magnet and vortex vigorously for 10 seconds. Re-engage the magnet and allow the beads to migrate to the magnet. Remove the wash fraction with a pipette and **combine with the saved depleted serum.**
9. Store the depleted fraction at or below -20 °C for long term storage.

Note: The sample may be concentrated and/or desalted prior to storage.

Elution of Proteins Bound to PureProteome Albumin/IgG Magnetic Beads (Optional)

The bound fraction of proteins may be analyzed by SDS-PAGE to ensure complete recovery of target protein(s) in the unbound depleted sample.

To elute the bound proteins, resuspend the beads 3 times using a minimum of 100 μ L of 200 mM Glycine-HCl, pH 2.0 for each elution. After adding the Glycine-HCl, disengage the magnet and vortex vigorously for 10 seconds. Re-engage the magnet and allow the beads to migrate to the magnet. Remove the eluted fraction with a pipette and save.


Alternatively, the bound proteins may be eluted in a 1X SDS-PAGE Reducing Sample Buffer (RSB). Resuspend the beads 1 to 3 times using 100 μ L of 1X RSB for each elution. After adding the RSB, disengage the magnet and vortex vigorously for 10 seconds. Remove tube and incubate at 70 °C for 10 minutes. Quickly place the tube back in the magnetic stand and allow the beads to migrate to the magnet. Immediately remove the eluted fraction with a pipette and save.

Note: The 12 kDa antibody ligand may be observed on the gel if the magnetic beads have been incubated in 1X RSB at 70 °C for 10 minutes.

Using Centrifugation for Concentration or Buffer Exchange (Optional)

Amicon Ultra-4 3K centrifugal filter devices are provided for rapid concentration and buffer exchange/desalting of the sample. Typical processing time is 20–30 minutes to reduce the volume of depleted sample to 50–150 μ L. A physical deadstop in the filter device prevents spinning to dryness and avoids potential sample loss. The concentrate is collected from the filter device sample reservoir using a pipettor. Concentration and buffer exchange/desalting of the depleted serum sample can be performed in the same device. A detailed user guide for the Amicon Ultra-4 filter device is provided.

1. Add the pooled depleted serum/wash fraction sample to the Amicon Ultra-4 filter device.
2. Assemble the filter device and centrifuge per instructions to concentrate the sample to the desired volume.
3. For applications that require buffer exchange or desalting, add 1.5 mL of the desired buffer and centrifuge again to concentrate the sample. Repeat this wash as required to ensure the desired exchange of buffer.



4. To recover the concentrated depleted fraction, insert a pipettor into the bottom of the filter device and withdraw the sample using a side-to-side sweeping motion.

Alternatively, concentration and buffer exchange may be performed using a different method, such as protein precipitation.

Disposal

Used material may be discharged into sewer or industrial waste water systems if allowed by local regulations. Otherwise, collect and dispose according to federal, state, and local regulations.

Material Safety Data Sheets (MSDS) are available on our web site. Go to www.millipore.com and enter your catalogue number in the search box.


Specifications

Matrix	Mixture of polymer-coated inorganic beads covalently coupled with recombinant Protein G and anti-albumin ligand.
Particle form	Spherical
Bead diameter	10 μ m (nominal)
Storage	2–8 °C. Do not freeze.
% Depletion	> 98% Albumin and IgG Typical values are ~99%

PureProteome Albumin/IgG Magnetic Beads are for research use only.

Performance


Figure 1. Removal of Human Serum Albumin (HSA) and IgG from Serum.



Coomassie blue-stained 2D PAGE gels of total human serum proteins (Panel A) and the fraction depleted of albumin and IgG (Panel B) using PureProteome Albumin/IgG Magnetic Beads. 100 μ g of total protein was resolved by isoelectric focusing (pH 4–7) in the first dimension and 8–16% SDS-PAGE in the second dimension.

Performance, continued


Figure 2. Human Serum (HS) Protein Analysis Pre- and Post-depletion.



Lane	1	2	3	4	5	6	7
Sample	HS	Unbound (depleted fraction)	Wash 1	Wash 2	Wash 3	Elution (bound fraction)	Molecular Weight Markers
Sample Volume	1 µL	4 µL	4 µL	4 µL	4 µL	4 µL	4 µL

Proteins were resolved on 4–12% SDS-PAGE and stained with Coomassie blue. HS (25 µL) was depleted of albumin and IgG following the protocol described for the PureProteome Albumin/IgG Magnetic Beads. The bound fraction was eluted from the magnetic beads using 100 µL additions of 200 mM Glycine-HCl, pH 2.0.

Figure 3. Depletion Efficiency of Human Serum Albumin (HSA) and IgG from Various Amounts of Human Serum.



Increasing amounts of human serum (25 µL, 50 µL, and 75 µL) were mixed with a fixed amount of PureProteome Albumin/IgG Magnetic Beads (950 µL of slurry or 170 µL of settled beads) and depleted as outlined in the protocol. The pre- and post-depleted HS samples were assayed by ELISA to calculate the percent depletion of both HSA and IgG.

Product Ordering Information

Description	Qty/Pk	Cat. No.
PureProteome Albumin/IgG Depletion Kit (contains magnetic beads, buffer concentrate, and Amicon Ultra-4 devices)	1	LSKMAGD12
PureProteome Albumin Magnetic Beads	10 mL	LSKMAGL10
PureProteome Magnetic Stand, 8-well	1	LSKMAGS08
PureProteome Magnetic Stand, 15 mL	1	LSKMAGS15
Amicon Ultra-4 3K Centrifugal Device	8 24 96	UFC800308 UFC800324 UFC800396

Additional Products for Downstream Analysis

Amicon Ultra-0.5 3K Centrifugal Device	8 24 96	UFC500308 UFC500324 UFC500396
ZipTip® SCX Pipette Tip, 0.6 µL strong cation resin	8 96	ZTSCXS008 ZTSCXS096
ZipTip C18 Pipette Tip, 0.6 µL C18 resin	8 96 960	ZTC18S008 ZTC18S096 ZTC18S960
ZipTip µC18 Pipette Tip, 0.2 µL C18 resin	8 96 960	ZTC18M008 ZTC18M096 ZTC18M960
ZipTip C4 Pipette Tip, 0.6 µL C4 resin	8 96 960	ZTC04S008 ZTC04S096 ZTC04S960

Notice

The information in this document is subject to change without notice and should not be construed as a commitment by Millipore Corporation or an affiliated corporation. Neither Millipore Corporation nor any of its affiliated corporations assumes responsibility for any errors that may appear in this document. This manual is believed to be complete and accurate at the time of publication. In no event shall Millipore Corporation be liable for incidental or consequential damages in connection with or arising from the use of this manual.

Technical Assistance

For more information, contact the Millipore office nearest you. In the U.S., call 1-800-MILLIPORE (1-800-645-5476). Outside the U.S., see your Millipore catalogue for the phone number of the office nearest you or go to our web site at www.millipore.com/offices for up-to-date worldwide contact information. You can also visit the tech service page on our web site at www.millipore.com/techservice.

Standard Product Warranty

The applicable Millipore warranty for the products listed in this publication may be found at www.millipore.com/ec/cp3/ terms (within the "Terms and Conditions of Sale" applicable to your purchase transaction).