



# Drummond Scientific Company

500 PARKWAY, BROOMALL, PENNSYLVANIA 19008-4203

PHONE : 610 - 353 - 0200

FAX 610 - 353 - 6204

www.drummondsci.com

May 19, 2025

## Certificate of Accuracy

We certify that the part number listed below is manufactured to provide a +/- 0.5% at full scale volumetric tolerance and comply with the quality standards of the Drummond Scientific Company. Measurement standards used during manufacture and testing are traceable to the National Institute of Standards and Technology (NIST) and are in compliance with the ASTM Standard E 672-87 Standard for Disposable Glass Micropipets.

### Part Number

2-000-020

### Description

20ul Drummond Calibrated Pipets

Alexis Piccirilli  
Quality Assurance