

# EU Declaration of Conformity



Product: Laboratory Equipment

File Reference: TFR910121\_002

Manufacturer: Antylia Scientific Ltd  
Beacon Road  
Stone, Staffordshire  
ST15 0SA  
United Kingdom

EU Representative

Antylia Scientific GmbH  
Futtererstraße 16  
97877 Wertheim  
Deutschland

This declaration of conformity is issued under the sole responsibility of the manufacturer

## Object of Declaration

**Cole-Parmer MP-250 series  
Melting Point Detectors**

The object of the declaration described above is in conformity with the relevant Union Harmonisation Legislation

**Electromagnetic Compatibility Directive**  
**Low Voltage Directive**  
**RoHS Directive**

**2014/30/EU**  
**2014/35/EU**  
**2011/65/EU (including CDD 2015/863)**

References to the relevant harmonised standards used or references to the other technical specifications in relation to which conformity is declared:

EN 61326-1:2013 Electrical equipment for the measurement, control and Laboratory use – Part 1: General requirements

EN 61010-1:2010 Safety requirements for electrical equipment for measurement, control and laboratory use – Part 1: General requirements  
/A1:2019/AC:2019-04

IEC / EN 61010-2-010:2014 Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-010: Particular requirements for laboratory equipment for the heating of materials

Signed for and on behalf of the above manufacturer

Original Year of Marking 2012  
Place of Issue Stone, UK  
Authorised Representative Martin Card  
Title EMEA Quality and Engineering Manager

Signature

Date 22 Sept 2021

Catalogue Number	Model	Description
03010-15	MP-250D	Mel-Temp melting point detector, fixed ramp rate
03010-51	MP-250D-F	Melting point detector, fixed ramp rate
03010-41	MP-250D-P	Melting point detector, programmable ramp rate
03010-59	MP-250D-PR	Melting point detector, beginning / ending recording