

CORNING GOSSELIN SAS  
123 Route de Caestre  
59190 - Borre  
FRANCE  
[www.corning.com/lifesciences](http://www.corning.com/lifesciences)  
Refer to website for regional contact information

Date of certificate : 2025 - 04 - 07

Page 1

### Product description

*Tube RoundBase2mL PS RedStraightPlu g Sterile 1250Bag 5000CS*

### Catalogue number

*TL40-04*

### Manufacturing date

*2025/03/01*

### Lot number

*921205*

*Lots : 921205*

### Quality Management System

Complies with the current version of the ISO9001 standard.

### Sterility Assurance Level

Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, "Sterilization of health care products - Radiation". Products meet a minimum Sterility Assurance Level (SAL) of 10<sup>-3</sup>.

### Expiration date

*2028/03/01*

### Expiration statement

*This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.*

### Performance testing (AQL)

*Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.*

This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release



Sandra Bianchi  
Plant Quality Manager, Borre  
[BianchiS@corning.com](mailto:BianchiS@corning.com)