Corning Incorporated
Life Sciences

Registered ISO 9001

Product description

Catalog Number HLC250-02
PDS reference PDS_Cbor_2337 Version _ 2
Product description Corning® Gosselin™ Water Sampling Square HDPE Bottle, 250 mL, Graduated, 20 mg/L Sodium Thiosulfate, 37 mm Red Tam

Sanitary Assurance Level
This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10-3

Expiration Statement
This product is designed to have an expiration date of 2 years from the date of manufacture regarding sterility.

Performance testing (AQL)
Inspection based on a statistically significant number of samples towards to internal that determine Assurance Quality Level.
Gravimetric dose of thiosulfate Na2S2O3.5H2: minimum dose 20mg/L

Primary body composition
HDPE

Secondary body composition
HDPE

Closure
37 mm Red Tamper-evident Cap with

Special feature
20 mg/L Sodium Thiosulfate

Packaging
1/Bag, 182/Case

DOMINIQUE DUTSCHER SAS

Regulation Statement

Bisphenol A, Phthalates, BSE/TSE, or Latex are not used in the manufacture of or intentionally incorporated into our products.
The raw materials for our products are food contact grade. Our suppliers guaranty the quality of their raw materials through a certificate of compliance.
However, our finished products have not been validated for food contact applications. They are designed for laboratory applications only.
The raw materials do not contain any substance that is listed on the current Article 59 list (so-called candidate list) of Regulation (EC) No 1907/2006 (REACH), above 0.1% (w/w).
The raw material do not contain Phthalates above 0.1 % (w/w).
The products contain less than 100 ppm of residual trace heavy metals (sum of Cd, Cr, Hg and Pb).

DOMINIQUE DUTSCHER SAS



Cedric Maes
Plant Quality Manager, Borre & Chorges
maesc@corning.com