SP SCIENCEWARE
Bel-Art · H-B Instrument · Wilmod-LabGlass
661 Route 23 South, Wayne, New Jersey 07470
Tel: 973-694-0500 * Fax: 973-694-7199
Internet Address: http://www.bel-art.com

CERTIFICATE OF ANALYSIS

This is to verify that all items noted below are in conformance with the Purchase Order Contract, drawings, specifications, and other applicable documentation; that all process certifications, material certifications, are on file at this or another Bel-Art facility and are available for review by the customer and or Government Representatives upon request.

Product #: 37876-0001 **Mfg. Date:** N/A **Lot #:** GENERAL

Description: Reusable Vaccu-Pette/96 Pipetting Device

Material: Polycarbonate

Dimensions: Dim. 5" x 3 3/8" x 1 7/16" High, Approximately .5ml per

Other: Autoclavable @ 121º C for 20 minutes @ 15 psi. Resins Meet FDA Regulation 21 CFR 177.1640
Manufactured: Bel-Art Products Pocomoke City MD, USA

Reviewed and Approved By:

Signature: 

Printed Name: Nancy K. Trogani
Title: Document Control Coordinator/Technical Support
Date: 11/14/2019