

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

**Product Description**

**Catalog Number:** 3383

**Product Description:** Corning® HTS Transwell, Feeder Reservoir, with lid, with Liquid Media Stabilizer

**Component Materials:**

- Reservoir & Lid - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.
- Liquid Media Stabilizer - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.

**Product Dimensions:**

- |                          |   |              |                    |   |          |
|--------------------------|---|--------------|--------------------|---|----------|
| Length of reservoir      | - | 5.030 in.    | Depth of Reservoir | - | .442 in. |
| Width of Reservoir       | - | 3.365 in.    | Height with Lid    | - | .626 in. |
| Tolerances of Dimensions | - | +/- .010 in. |                    |   |          |

**Sterilization:**

The product is irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*.

Sterility Assurance Level: SAL 10<sup>-3</sup>

**Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual examination of the product.

Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 4