

**Corning Incorporated
Life Sciences**

Registered
ISO 9001:2008

Product Description

Catalog Number: 3473

Product Description: Costar® 24-well Ultra Low Attachment multiple well plate, flat bottom, with lid

Component Materials:

- Plate/Lid - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.
Surface - Proprietary hydrogel.

Product Dimensions:

Length of Plate	-	5.030 in.	Diameter of Well @ top	-	.640 in.
Width of Plate	-	3.365 in.	Diameter of Well @ bottom	-	.615 in.
Depth of Well	-	.685 in.	Height with Lid	-	.891 in.
Tolerances of Dimensions	-	+/- .050 in.	Growth Area	-	1.9 cm ²

Sterilization:

The product is irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*.
Sterility Assurance Level: SAL 10⁻³

Pyrogens:

The product has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72:2002/(R)2010 *Bacterial Endotoxins - Test methodologies, routine monitoring, and alternative to batch testing*.
Results: ≤ 0.1 EU/mL (≤ 4EU/device)

Surface Characterization:

Surface is characterized to be hydrophilic and neutrally charged, composed of a covalently bound hydrogel layer that is biologically inert and non-degradable. This surface composition has been optimized for cell attachment inhibition.

Cell Attachment and Growth Characteristics:

The product has been tested for the attribute of cell attachment inhibition utilizing an attachment- dependent mammalian cell line in a serum supplemented media. Cell attachment inhibition must be ≥ 98% as compared to a standard tissue culture treated surface.

Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:

This product is manufactured and packaged with animal free materials.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual and microscopic examination of the product.

Packaging:

Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 15