Corning Incorporated Life Sciences

Registered ISO 9001:2008

Product Description

Catalog Number: 3896

Product Description: Costar® 96-well assay plate, v-bottom without lid

Component Materials:

Virgin Polystyrene, meets USP, Class VI requirements for plastic containers and

closures.

Product Dimensions:

.270 in. Length of Plate 5.035 in. Diameter of Well @ top Width of Plate Diameter of Well @ 3.370 in. .250 in. bottom Depth to the bottom of .420 in. Height .560 in.

well

Working Volume per well -Tolerances of +/-.010 in. 0.1 - 0.2 mL

Dimensions

Sterilization:

The product has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization. Sterility Assurance Level: SAL 10⁻³

Surface Characterization:

Surface is characterized to be hydrophobic and non-charged for solution based assay.

Optical Characteristics:

The product is made of clear polystyrene to allow transparency.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures. Visual Attributes: Visual and microscopic examination of the product.

Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct

product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 10