

Digital Hand-held "Pocket" High Fructose Corn syrup (HFCS-55) Refractometer PAL-17S



PAL-17S is suitable to measure the concentration of High Fructose Corn Syrup (HFCS-55) water solution.

Specifications

Model	PAL-17S
Cat.No.	4417
Scale	High Fructose Corn Syrup (HFCS-55)(g/100g)
Range	0.0 to 85.0%(g/100g)
Accuracy	±0.2%
Temperature compensation	10 to 40° C

Accessories

- PAL-CASE : RE-39409
- JORDAN for PAL(Silicon Cover) : RE-39414
- STRAP : RE-39410
- Battery Case Cover With O-Ring : RE-38163
- USB Battery Charger : RE-34801

Patent for Invention Registration No. ZL200310103015.2 (China)
Patent for Design Registration No. ZL03303431.1 (China), 089244 (Taiwan)
Patent Granted in countries around the world.