

Technical Compliance Statement


Ref No. EM971688

For the following equipment

Product : Genius Dry Bath Incubator
Model Number : (1)MD-01N (2)MD-02N
Applicant : Major Science Co., Ltd.
Manufacturer : Major Science Co., Ltd.
Standards :
EN 61326-1/2006
EN 61326-2-2/2006
Emission: EN 55011/2007, Class B
(CISPR 11/2003 +A1/2004, Class B)
EN 61000-3-2/2006 and EN 61000-3-3/1995+ A1/2001+ A2/2005
Immunity: EN 61000-4-2/1995+A1/1998+A2/2001, EN 61000-4-3/2006,
EN 61000-4-4/2004, EN 61000-4-5/2006, EN 61000-4-6/1996+A1/2001,
EN 61000-4-11/2004

We hereby certify that the above product has been tested by us with the listed standards and found in compliance with the council EMC directive 2004/108/EC. It is possible to use CE marking to demonstrate the compliance with this EMC Directive. The test data & results are issued on the EMC test report no. EM-E970304.

Signature



Leon Liu/General Deputy Manager
Date: Oct. 28, 2008

Test Laboratory:
AUDIX Technology Corporation, EMC Department
DAR Reg. No.: DAT-P-145/03-01
Web Site: www.audixtech.com

The statement is based on a single evaluation of one sample of the above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.



Reg. No: DAT-P-145/03-01