

Nolato Treff Certificate of Conformity

CoC Certificate according to DIN EN 10204 2.1

	Article:	PCR Single Tube PP clear 0.5 ml
	Article Number:	96.10385.9.01
	Purchase Number:	PO x
	Lot- Number:	5x
	Quantity:	10'000 pcs

Statement

Nolato Treff AG confirms that TreffLab consumables and products marked with the CleanRoom[®] Pure quality symbol have been manufactured in microbiologically controlled clean rooms class 8, ISO 14644-1. To avoid direct human contact, employees are fully protective dressed and masked.

Quality Parameters

Quality standard for original sealed packages

- Free of RNase (< 5x10E-9 Kunitz units)
- Free of DNase (< 5x10E-6 Kunitz units)
- Free of DNA (< 2 pg DNA)
- Free of endotoxins (< 10 copies)
- Free of PCR inhibitors (< 0.001 EU/ml)
- Free of R and ATP (< 30 RLU)

Analyses are performed by an independent Swiss laboratory for molecular biological & microbiological analytics, accredited ISO/IEC17025 for real-time PCR.

Date: 28 November, 2023
Place: CH-9113 Degersheim
Nolato Treff AG

Customer Service