

Edition Date: Revision Date : Revised by:

2016-09-15 2017-11-03 Annette Roy, QA Coordinator

Product Name: Micrewtube[®] cap Catalogue No.: T340_OSFTTP This document replaces any previous version 1. **Product Description:** Tamper Evident Micrewtube[®] Screw Cap: Disposable, non-sterile tamper evident flattop cap with Silicone washer seal for Micrewtube[®]. Tubes sold separately. > Available colors: • Natural, Blue, Green, Lilac, Orange, Red, Yellow, White, Brown > Compatible with: • T341-XTTP & T341-XTPRTP Tamper Evident Micrewtube® Series 2. Packaging: Case: 1 000 units per package 3. **Product Specifications:** ➤ Material: • Cap: Polypropylene • Washer Seal: Silicone Gamma sterilization resistant; Temperature range: -196°C to +121°C; Certified RNase, DNase, Pyrogen and DNA-free; Autoclavable at 121°C, for up to 30 minutes. 4. **Standards and Conformity:** Sampling and inspection procedures

- > ISO 2859-1:
- > FDA:
- > USP:
- > CONEG / RoHS:

Resin conforms to FDA 21 CFR 177.1520 Resin conforms to USP Class VI

Plastic is in conformity to REACH standards

Plastics and colorants are in conformity with CONEG / RoHS standards for heavy metals

REACH (SVHC):



Edition Date: Revision Date : Revised by: 2016-09-15 2017-11-03 Annette Roy, QA Coordinator

5. Quality Assurance:

- > Clear, no presence of contamination in plastic
- Visual attributes
- Closure verification
- Leak proof testing in vacuum at 71.3cm Hg
- Gas phase of Liquid Nitrogen resistance

6. Traceability:

Lot No. Composition: 8 or 9 digits

> The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > For use in automated equipment, follow the equipment manufacturer's instructions
- > Should be used only in the gas phase of Liquid Nitrogen

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.