Edition Date: 2004-02-09 Revision Date: 2017-10-03

Revised by: Annette Roy, QA Coordinator

Catalogue No.: C572-2 **Product Name: Sample Tube**

This document replaces any previous version

1. **Product Description:**

> 30ml Sample Tube with Cap: Disposable, non-sterile and conical bottom tube with self-standing base

2. Packaging:

Case: 5 packs of 100 units / 500 units per case

3. **Product Specifications:**

Material:

• Tube: Polypropylene • Cap: Polyethylene • Cap Color: white

- > Gamma radiation resistant
- > Temperature range: -90°C to +100°C
- Tubes have molded-in graduations from 10ml to 30ml with 5ml increments
- ➤ Cap configuration allows insertion of a CapinsertTM (T345 Series) for color-coding

Standards and Conformity: 4.

> ISO 2859-1: Sampling and inspection procedures ➤ FDA: Resins conform to FDA 21 CFR 177.1520

➤ USP: Resins conform to USP Class VI

> CONEG / RoHS: Plastics and colorants are in conformity with

CONEG / RoHS standards for heavy metals

> REACH (SVHC): Plastics are in conformity to REACH standards

> LATEX: Materials are Latex Free ➤ BSE / TSE: Materials are BSE / TSE Free

➤ CE: Product is CE marked

5. **Quality Assurance:**

- Clear, no presence of contamination in plastic
- Visual attributes
- Volume measurements
- Closure verification
- Gas phase of Liquid Nitrogen resistance

Page 1 / 2 Issue Date: 2017-06-15

Edition Date: 2004-02-09 **Revision Date:** 2017-10-03

Revised by: Annette Roy, QA Coordinator

6. Traceability:

Lot No. Composition: 8 or 9 digits

> The lot number can be found in one or all of these locations:

- 1. On exterior case label
- 2. On label inserted inside the master case
- 3. On the inner bag

7. Storage Conditions:

- > Store at room temperature in normal warehouse conditions
- > Avoid temperature variations and humidity
- Protect from any possible contamination
- Protect from any damage to the packaging

8. Recommended Use:

- > Verify proper cap closure when using biohazard material and / or chemical reagents
- > Follow chemical resistance chart recommendations
- > Follow the equipment manufacturer's instructions
- > Should be used only in the gas phase of Liquid Nitrogen

OWNERSHIP OF MATERIALS:

Materials and information contained in this document are **Simport Scientific**'s copyrighted and are protected by worldwide copyright laws and treaty provisions. They may not be copied, reproduced, modified, published, uploaded, posted, transmitted or distributed in any way.

Document Template: FRM-106 Rev01 Technical Data Sheet

Issue Date: 2017-06-15