## **Technical Service Consultants Ltd**

Microbiology House, Fir Street, Heywood, Lancashire OL10 1NW, UK T : +44 (0)1706 620600 F : +44 (0)1706 620445

E:sales@tscswabs.co.uk W:www.tscswabs.co.uk

## **Product Specification Sheet**

## TS/6-61:250

PreMoist Swab - Polystyrene shaft/Viscose Swab, pre-moistened in TSC Neutralising Buffer sealed in tube

Product Code	TS/6-61:250
Description	PreMoist Swab - Polystyrene shaft/Viscose Swab, premoistened in TSC Neutralising Buffer sealed in tube
Intended Use	Sampling of surfaces for environmental and hygiene monitoring, infection control and cleaning validation
Components	Polypropylene tube, Blue HDPE Standard Cap, White Polystyrene/Viscose swab, Label, Non toxic glue.
Contents	De-ionised Water, Tween 80, Lecithin, Sodium Thiosulphate, Sodium Thioglycollate, Sodium sulfite, Buffers to maintain. pH - 7.2 ± 0.2.
Method of Sterilisation	Sterilised by gamma irradiation at dosage of 25-53.4kGy
Shelf Life	12 months from date of manufacture
Storage Conditions	4-25°C. Store in dry and ventilated area; protect from light, heat and moisture in tightly closed original packaging
Precautions and Limitations	Analyse as soon as possible due to the small volume, so to minimise sample dryout and organism loss. Refrigeration will stabilise recovery of some organisms. Not suitable for most fastidious organism unless analysed immediately. Due to product characteris
Pack Size	250 in a box
Other	Quality control testing details on final product: i) Validation of irradiation sterility ii) Product appearance.

This specification (and formulation where applicable) are intended for guidance only and may be adjusted to meet product defined criteria.

You should always refer to a professional with experience in your intended application if in any doubt.

Issue #: 2 Date: 24.04.2020 ProductSpec\_TS6-61:250