Technical Service Consultants Ltd

Microbiology House, Fir Street, Heywood, Lancashire OL10 1NW, UK T: +44 (0)1706 620600 $\,$ F: +44 (0)1706 620445

E: sales@tscswabs.co.uk W: www.tscswabs.co.uk

Product Specification Sheet

TS/6-62:100.

PreMoist Swab - Polystyrene shaft/Viscose Swab, pre-moistened in de-ionised water sealed in tube

Draduct Cada	TC/C C0:100
Product Code	TS/6-62:100.
Description	PreMoist Swab - Polystyrene shaft/Viscose Swab, pre-
	moistened in de-ionised water sealed in tube
Intended Use	Environmental sampling, Not suitable for prolonged transport,
	cleaning validation or hygiene monitoring unless samples are
	processed on growth media with neutralisers immediately.
Components	Polypropylene Tube, Blue HDPE Standard Cap,
Components	Polystyrene/viscose tip swab, Label, hot melt glue.
Contents	De-ionised Water - 1 Litre
Contents	De-ionised water - 1 Litte
Mathada Gold Pradica	Other History and Control of the Con
Method of Sterilisation	Sterilised by gamma irradiation at dosage of 25-53.4kGy
Shelf Life	12 months from date of manufacture
Storage Conditions	4-25°C. Store in dry and ventilated area; protect from light,
	heat and moisture in tightly closed original packaging
Precautions and	Analyse as soon as possible due to the small volume, so to
Limitations	minimise sample dryout and organism loss. Refrigeration will
	stabilise recovery of some organisms. Not suitable for most
	fastidious organism unless analysed immediately.
Pack Size	100 in a box
I don Oize	
Other	Quality control testing details on final product: i) Validation of
Other	, , , , , , , , , , , , , , , , , , , ,
	irradiation sterility ii) Product appearance.

This specification (and formulation where applicable) are intended for guidance only and may be adjusted to meet product defined criteria.

You should always refer to a professional with experience in your intended application if in any doubt.

Issue #: 2 Date: 24.04.2020 ProductSpec_TS6-62:100.