

## ANNEX

### EU TYPE EXAMINATION CERTIFICATE Nr. 032/2023/0488

#### 1. Applicant

Ansell Healthcare Europe NV/SA  
Riverside Business Park  
Block J Boulevard International 55  
1070 Brussels  
Belgium

#### 2. Description

EN ISO 374-1/Type B



K P T

EN ISO 374-5:2016



VIRUS

EN 421:2010



#### 3. Materials and accessories

##### Gloves

- MicroFlex® 93-833



## 4. Technical documentation

### Summary test results

EN ISO 21420:2020

Gloves

**MicroFlex® 93-833**

Method	Description	Result	Class
EN 14362-1	AZO dyes for colored gloves	PASS	
ISO 3071:2020	pH - textile	PASS	
ISO 16190:2021	Polycyclic aromatic hydrocarbons (PAHs)	PASS	
EN 16778:2016	Dimethylformamide (DMFa)	/	/
EN ISO 21420:2021 §6.1	Length	PASS	
1149-1 / 1149-3 or EN 16350	Electrostatic properties	/	/
EN ISO 21420:2020 §6.2	Dexterity	PASS	Level 5
ISO 11092:2014 textile	Water vapour transmission	/	/

EN ISO 374-1:2016+A1:2018

Gloves

**MicroFlex® 93-833**

Method	Description	Result	Class
EN 374-2:2014	Penetration	PASS	
EN 16523-1:2015+A1:2018	Permeation K	PASS	Level 6
EN 16523-1:2015+A1:2018	Permeation P	PASS	Level 4
EN 16523-1:2015+A1:2018	Permeation T	PASS	Level 6
EN ISO 374-4:2019	Degradation K	PASS	
EN ISO 374-4:2019	Degradation P	PASS	
EN ISO 374-4:2019	Degradation T	PASS	
Type of glove		PASS	Type B

EN ISO 374-5:2016

Gloves

**MicroFlex® 93-833**

Method	Description	Result	Class
EN 374-2:2014	Penetration	PASS	
ISO 16604:2004 Proc. B	Protection against viruses	PASS	

EN 421:2010

Gloves

**MicroFlex® 93-833**

Method	Description	Result	Class
EN 420 or EN ISO 21420	Dexterity	PASS	Level 5
EN 16523-1:2015+A1:2018	Permeation K	PASS	Level 6
EN 16523-1:2015+A1:2018	Permeation P	PASS	Level 4
EN 16523-1:2015+A1:2018	Permeation T	PASS	Level 6
EN ISO 374-2	Penetration	PASS	
EN 388	Mechanical requirements	Dexterity is the most important parameter	/
EN 421 4.3 method 5.1	Attenuation efficiency and uniformity of distribution	/	/

## Description/Picture of article

Article      **The Gloves MicroFlex® 93-833**



The above picture is a general picture of the article. Possible variations of the above article can be present in the technical file.

Note :

Any modification in material, design, or other technical features must be brought to the attention of the Notified Body.