

Thermo Scientific Finnpiquette Performance Specifications

Salla Ruosaari, Product Manager, Thermo Fisher Scientific, Vantaa, Finland.

Key words

- **ISO 8655** - Calibration standard for the piston operated volumetric devices
- **Factory Specifications** - Often more stringent requirements than the above ISO 8655, these are set by manufacturers in the development phase of a new product
- **Systematic error** - The difference between the dispensed volume and the selected volume of a pipette
- **Random error** - Error referring to the repeatability of pipetting
- **Pipetting Systems** - In liquid handling, pipettes and tips always form a pipetting system. Using a pipette with the pipette tip it was designed, developed and manufactured to fit to allow for ideal pipetting performance
- **ISO 9001** - Standard for Quality management systems
- **ISO 13485** - Quality management standard for medical devices

ISO 8655 Specifications

The ISO 8655 standard forms a guideline for industry and independent service companies concerning the entire calibration process of piston operated volumetric devices. The acceptance limits stated in the ISO standard are designed to be used by manufacturers, service companies as well as end users. The standard criteria are agreed to be sufficient for the use with the devices described therein.

The standard gives the maximum permissible limits for systematic and random error, also referred to as inaccuracy and imprecision, respectively. Systematic error is the difference between the dispensed volume and the selected volume of a pipette. The error is repeated in each measurement taken but can be eliminated if the error source, such as imperfect calibration or a change in the environment where the testing is performed, is identified. Random error, on the other hand, refers to the repeatability of the pipetting. It is caused by random factors including features of the pipette, laboratory practice and user experience.



Factory Specifications

All manufacturers, Thermo Fisher Scientific included, use their own factory performance specifications to describe the actual functionality of their pipettes. As the factory specifications are tighter than those required by the ISO 8655 standard, this is an acceptable practice in the industry. These specifications are provided for pipetting systems, which comprises both pipette and tip. Performance achievements may be different if using non-system tips. The Finnpiquette performance has been tested using their system tip which are Thermo Scientific Finntips.

Optimal Pipetting Performance with the Thermo Scientific Systems

The main difference between the factory and ISO 8655 specifications is that the factory error limits are testing volume dependent for variable volume pipettes. The ISO 8655 standard, on the other hand, gives the same error limits for the whole volume range of each pipette. The difference is especially noticeable in the lower volume error limits ensured by Thermo Scientific. For the mid and minimum volumes, the factory specifications are remarkably tighter than what is required by the ISO standard. With more stringent specifications, pipetting performance is guaranteed with the Finnpiettes.

To demonstrate the difference, ISO 8655 specifies a systematic error of 8 µL for a 1000 µL pipette regardless of the selected volume. Thermo Scientific F1 and F2 100-1000 µL pipettes are however given a systematic error of 1 µL at minimum and 6 µL at maximum volume. Thus, when pipetting a sample of 100 µL, the maximum permissible error according to Thermo Scientific factory specifications will be only 1% but 8% according to ISO 8655 (Table 1). The same applies for all Finnpiette models as demonstrated in Table 2

Quality matters

Your tip choice has a big impact on the pipetting performance, which is why Finnpiettes are designed for use with the Finntips. Both Finnpiettes and Finntips are manufactured from high quality raw materials, tested thoroughly and certified according to strict quality standards. Specifically, Thermo Fisher Scientific Oy is a company certified to the standards ISO 9001 and ISO 13485, taking quality matters seriously. Further, the pipetting systems comprising Finnpiettes and Finntips are CE-marked according to IVD Directive in Europe and intended for use in in-vitro diagnostics for transferring of liquids.

For further information on calibration of pipettes, refer to The Good Laboratory Pipetting Guide.

Volume	Thermo Scientific Factory Specifications		ISO 8655 Specifications	
	Inaccuracy µl (%)	Imprecision µl (%)	Inaccuracy µl (%)	Imprecision µl (%)
1000	±6,0 (±0,6%)	2,0 (0,2%)	±8,0 (±0,8%)	3,0 (0,3%)
500	±4,0 (±0,8%)	1,5 (0,3%)	±8,0 (±1,6%)	3,0 (0,6%)
100	±1,0 (±1,0%)	0,6 (0,6%)	±8,0 (±8,0%)	3,0 (3,0%)

Table 1. Comparison of the Thermo Scientific Factory Specifications, and ISO 8655 limits for Finnpiette F1 and F2 100-1000 µl models. Notice how remarkably the mid and min volume error limits differ between the Thermo Scientific Factory Specifications and ISO 8655.

Thermo Scientific Finnpiquette F3 Pipettes*

Finnpiquette F3 Variable Volume Single Channel Pipettes



Cat. No.	Range	Volume (µL)	Thermo Scientific Factory Calibration Specifications		ISO 8655 Calibration Specifications	
			Inaccuracy (µL)	Imprecision (µL)	Inaccuracy (µL)	Imprecision (µL)
4640000	1-10 µL	10	±0,100	0,050	±0,120	0,080
		5	±0,100	0,050	±0,120	0,080
		1	±0,025	0,020	±0,120	0,080
4640010	1-10 µL	10	±0,100	0,080	±0,120	0,080
		5	±0,100	0,065	±0,120	0,080
		1	±0,035	0,030	±0,120	0,080
4640020	2-20 µL	20	±0,20	0,08	±0,20	0,10
		10	±0,15	0,07	±0,20	0,10
		2	±0,06	0,05	±0,20	0,10
4640030	2-20 µL	20	±0,20	0,08	±0,20	0,10
		10	±0,15	0,07	±0,20	0,10
		2	±0,06	0,05	±0,20	0,10
4640080	5-50 µL	50	±0,30	0,15	±0,50	0,20
		25	±0,25	0,13	±0,50	0,20
		5	±0,15	0,125	±0,50	0,20
4640090	5-50 µL	50	±0,30	0,15	±0,50	0,20
		25	±0,25	0,13	±0,50	0,20
		5	±0,15	0,125	±0,50	0,20
4640040	10-100 µL	100	±0,80	0,20	±0,80	0,30
		50	±0,60	0,20	±0,80	0,30
		10	±0,30	0,10	±0,80	0,30
4640050	20-200 µL	200	±1,2	0,6	±1,60	0,60
		100	±1,0	0,5	±1,60	0,60
		20	±0,6	0,3	±1,60	0,60
4640060	100-1000 µL	1000	±5,0	2,0	±8,0	3,0
		500	±5,0	2,0	±8,0	3,0
		100	±1,5	0,6	±8,0	3,0
4640100	0,5-5 mL	5000	±25,0	10,0	±40,0	15,0
		2500	±17,5	7,5	±40,0	15,0
		500	±10,0	4,0	±40,0	15,0
4640070	1-10 mL	10000	±50,0	20,0	±60,0	30,0
		5000	±40,0	15,0	±60,0	30,0
		1000	±20,0	8,0	±60,0	30,0

Finnpiquette F3 Fixed Volume Single Channel Pipettes



Cat. No.	Range	Volume (µL)	Thermo Scientific Factory Calibration Specifications		ISO 8655 Calibration Specifications	
			Inaccuracy (µL)	Imprecision (µL)	Inaccuracy (µL)	Imprecision (µL)
4650000	1 µL	1	±0,040	0,040	±0,050	0,050
4650010	5 µL	5	±0,070	0,070	±0,125	0,075
4650020	10 µL	10	±0,090	0,080	±0,120	0,080
4650090	20 µL	20	±0,14	0,10	±0,20	0,10
4650030	25 µL	25	±0,15	0,125	±0,50	0,20
4650040	50 µL	50	±0,30	0,20	±0,50	0,20
4650050	100 µL	100	±0,40	0,30	±0,80	0,30
4650100	200 µL	200	±0,80	0,60	±1,60	0,60
4650060	250 µL	250	±1,0	0,8	±4,0	1,5
4650070	500 µL	500	±1,5	1,5	±4,0	1,5
4650080	1000 µL	1000	±3,0	3,0	±8,0	3,0

* Please note: Availability may vary by country.