

User Guide

Millex®-AP₅₀ Prefilter, Cat. No. SLAP05010

Introduction

The Millex®-AP₅₀ filter unit is recommended for use as a prefilter for tissue culture media, prior to final filtration with a Sterivex™ filter unit. The Millex®-AP₅₀ filter extends the life of the Sterivex™ filter unit by filtering out large particles that can clog the smaller pore size filter in the Sterivex™ unit. The Millex®-AP₅₀ filter can be connected directly to the Sterivex™ filter using the supplied Millex®-Sterivex™ connector.

CAUTION: The Millex®-AP₅₀ filter is for prefiltration only. It may release fibers into the filtrate, so a final membrane filter should always be installed downstream of the prefilter.

Directions for Use

1. Attach the barbed hose inlet of the Millex®-AP₅₀ prefilter to the outlet side of the peristaltic pump tubing and immerse the inlet end of the tubing into the solution to be filtered.
2. Insert the non-Luer-Lok™ end of the Millex®-Sterivex™ connector into the open end of the Millex®-AP₅₀ barbed hose outlet.
3. Aseptically remove the sterile Sterivex™ filter unit from the blister pack and connect it to the Luer-Lok™ end of the Millex®-Sterivex™ connector.
4. Place a receiving container under the Sterivex™ filter unit outlet and turn on the pump.

NOTE: The Sterivex™ filter unit has a hydrophobic 0.1 µm polyvinylidene fluoride (PVDF) vent filter that allows air trapped in the system to be automatically expelled.

Specifications

Materials	Filter: AP20 borosilicate glass microfiber Housing: Polypropylene
Dimensions	Length: 71 mm (2.8 in.) Diameter: 62 mm (2.4 in.) Filtration area: 19.6 cm ² (3.0 in ²)
Connections	Millex®-AP ₅₀ filter Inlet and outlet: stepped hose barb (for 5–12 mm [0.2–0.5 in.] ID tubing) with female Luer-slip interior Millex®-Sterivex™ connector Inlet: Male Luer-slip Outlet: Male Luer-Lok™
Temperature limit	45 °C (113 °F)
Pressure limit at 25 °C	2.1 bar (30 psi)
Hold-up volume	2.2 mL
Water flow rate	≥ 650 mL/min at 0.7 bar (10 psi), 25 °C

Ordering Information

Product Description	Qty/pk	Cat. No.
Millex®-AP ₅₀ filter unit	10	SLAP05010
Sterivex™-GV filter unit without filling bell	15	SVGV01015
Sterivex™-GV filter unit without filling bell	50	SVGV010RS
Sterivex™-GV filter unit with filling bell	10	SVGVB1010
Sterivex™-GV filter unit with Luer-Lok™ outlet	15	SVGVL10RC
Sterivex™-HV filter unit without filling bell	15	SVHV01015
Sterivex™-HV filter unit without filling bell	50	SVHV010RS
Sterivex™-HV filter unit with filling bell	10	SVHVB1010
Sterivex™-HV filter unit with Luer-Lok™ outlet	15	SVHVL10RC
Sterivex™-GP filter unit without filling bell	15	SVGP01015
Sterivex™-GP filter unit without filling bell	50	SVGP01050
Sterivex™-GP filter unit with filling bell	10	SVGPB1010
Sterivex™-GP filter unit with Luer-Lok™ outlet	15	SVGPL10RC

Notice

The information in this document is subject to change without notice and should not be construed as a commitment by Millipore S.A.S. ("Millipore") or an affiliate. Neither Millipore S.A.S. nor any of its affiliates assumes responsibility for any errors that may appear in this document.

Technical Assistance

For more information, contact the office nearest you. In the U.S., call 1-800-MILLIPORE (1-800-645-5476). Outside the U.S., go to our web site at www.millipore.com/offices for up-to-date worldwide contact information. You can also visit the tech service page on our web site at www.millipore.com/techservice.

Standard Warranty

The applicable warranty for the products listed in this publication may be found at www.millipore.com/terms (within the "Terms and Conditions of Sale" applicable to your purchase transaction).