



Benchmark Scientific Inc.
PO Box 709
Edison, NJ 08818

EU DECLARATION OF CONFORMITY

Benchmark Scientific, Inc. hereby declares that the device(s) described below are in conformity with the EC Directives listed here-in as of the date of this declaration. In the event of unauthorized modification of any of the devices listed below, this declaration becomes invalid.

Device Name: MultiTherm and Incu-Mixer Microplate Shaking Incubators

Device Model Number(s): H6002 (-E), H5000-H (-E)
H6004 (-E) H5000-HC (-E)

*Including private label versions of these devices.

Relevant EC Directives: Low Voltage Directive 2014/35/EU
EMC Directive 2014/30/EU
RoHS2 2011/65/EU

Harmonized Standards: EN61326-1:2013
EN 61010-2-010:2014
EN61010-1:2010

Manufacturer's Name: Benchmark Scientific, Inc.
PO Box 709
Edison, NJ 08818

Date: January 1st, 2019

Vice President