

Benchmark Scientific Inc. PO Box 709 Edison, NJ 08818

## **EU DECLARATION OF CONFORMITY**

Benchmark Scientific, Inc. hereby declares that the device(s) described below are in conformity with the EC Directives listed here-in as of the date of this declaration. In the event of unauthorized modification of any of the devices listed below, this declaration becomes invalid.

| Device Name:            | Plate and Tube Shakers/Vortexers                                   |                                  |           |
|-------------------------|--|----------------------------------|-----------|
| Device Model Number(s): | BV1010 (-E)  | BT1502 (-E)                      |           |
|                         | *Includes all color var  | iations and private label produc | ets       |
| Relevant EC Directives: | Low Voltage Directive<br>EMC Directive 2014/30<br>RoHS2 2011/65/EU |                                  |           |
| Harmonized Standards:   | EN61326-1:2013<br>EN61010-1:2010                                   |                                  |           |
| Manufacturer's Name:    | Benchmark Scientific, 1<br>PO Box 709<br>Edison, NJ 08818          | Inc                              |           |
| Date:                   | January 02, 2019   |                                  |           |
|                         |  | 10                               | ny Dermen |

Marketing Manager