

Benchmark Scientific Inc. PO Box 709 Edison, NJ 08818



Benchmark Scientific, Inc. hereby declares that the device(s) described below are in conformity with the EC Directives listed here-in as of the date of this declaration. In the event of unauthorized modification of any of the devices listed below, this declaration becomes invalid.

Device Name: Handheld Homogenizers

Device Model Number(s): D1000 (-E)

Relevant EC Directives: Low Voltage Directive 2014/35/EU

EMC Directive 2014/30/EU

RoHS2 2011/65/EU

Harmonized Standards: EN61326-1:2013

EN61010-1:2010

Manufacturer's Name: Benchmark Scientific, Inc

PO Box 709 Edison, NJ 08818

Date: January 02, 2019

Marketing Manager

Phone: (908) 769-5555 Fax: (732) 313-7007 Email: info@BenchmarkScientific.com