

Benchmark Scientific Inc.
PO Box 709
Edison, NJ 08818

EU DECLARATION OF CONFORMITY

Benchmark Scientific, Inc. hereby declares that the device(s) described below are in conformity with the EC Directives listed here-in as of the date of this declaration. In the event of unauthorized modification of any of the devices listed below, this declaration becomes invalid.

Device Name: Accuris MyGel Mini Electrophoresis system

Device Model Number(s): E1101-E, E2102-F-E and private label versions thereof

Relevant EC Directives: Low Voltage Directive 2014/35/EU
EMC Directive 2014/30/EU
RoHS2 2011/65/EU

Harmonized Standards: EN61326-1:2013
EN61010-1:2010

Manufacturer's Name: Benchmark Scientific, Inc
PO Box 709
Edison, NJ 08818

Date: January 02, 2019



Marketing Manager