

Benchmark Scientific Inc.
PO Box 709
Edison, NJ 08818

DECLARATION OF CONFORMITY

Benchmark Scientific, Inc. hereby declares that the device(s) described below are in conformity with the EC Directives listed here-in as of the date of this declaration. In the event of unauthorized modification of any of the devices listed below, this declaration becomes invalid.

Device Name: Accuris Analytical, Precision and Compact Electronic Balances

Device Model Number(s): W4000-100, W4000-500, W3100-120-E, W3100-210-E, W3100A-120-E, W3100AA-210-E, W3200-120-E, W3200-320-E, W3200-500-E, W3200-1200-E, W3200-3200-E, W3200-5K-E, W3300-120-E, W3300-300-E, W3300-500-E, W3300-1200-E, W3300-5K-E, W3300-10K-E and Private label versions of these devices.

Relevant EC Directives: EC EMC directive 2014/30/EU with amendments.

Harmonized Standards: EN 61326-1:2013
EN 61326-2-2:2013(CLASS B)
EN 61000-3-2:2014
EN 61000-3-3:2013

Remarks:

The CE markings as shown below can be affixed on the product after preparation of necessary conformity documentation, as stipulated in article 10 of the Council Directive **93/68/EEC**.

Manufacturer's Name: Benchmark Scientific, Inc
PO Box 709
Edison, NJ 08818

Date: June 20, 2016



Marketing Manager