

Benchmark Scientific Inc.
PO Box 709
Edison, NJ 08818

DECLARATION OF CONFORMITY

Benchmark Scientific, Inc. hereby declares that the device(s) described below are in conformity with the EC Directives listed here-in as of the date of this declaration. In the event of unauthorized modification of any of the devices listed below, this declaration becomes invalid.

Device Name: Accuris™ SmartDoc™ Gel Visualization, Smart Phone Imaging System

Device Model Number(s): E5001-SDB, E5001-SDB-E

Relevant EC Directives: Directive 2006-95-EC
EMC Directive 2004/108/EC (EN 61326-1)

Applicable Standards: EN 61326-1:2013 EMC
EN 61010-1:2010 +A1:2019 LVD

Manufacturer's Name: Benchmark Scientific, Inc.
PO Box 709
Edison, NJ 08818

Date: September 9, 2019



Vice President