



Benchmark Scientific Inc.  
PO Box 709  
Edison, NJ 08818

## EU DECLARATION OF CONFORMITY

Benchmark Scientific, Inc. hereby declares that the device(s) described below are in conformity with the EC Directives listed here-in as of the date of this declaration. In the event of unauthorized modification of any of the devices listed below, this declaration becomes invalid.

**Device Name:** Thermal Cyclers

**Device Model Number(s):** T5000-96 (-E), T5005-3205 (-E)

\*Including private label versions of these devices.

**Relevant EC Directives:** Low Voltage Directive 2014/35/EU  
EMC Directive 2014/30/EU  
RoHS2 2011/65/EU  
RoHS2 2015/863/EU

**Harmonized Standards:** EN61326-1:2013  
EN 61010-2-010:2014  
EN61010-1:2010

**Manufacturer's Name:** Benchmark Scientific, Inc.  
PO Box 709  
Edison, NJ 08818

**Date:** January 1st, 2019



\_\_\_\_\_  
Vice President