

Benchmark Scientific Inc.
PO Box 709
Edison, NJ 08818

EU DECLARATION OF CONFORMITY

Benchmark Scientific, Inc. hereby declares that the device(s) described below are in conformity with the EC Directives listed here-in as of the date of this declaration. In the event of unauthorized modification of any of the devices listed below, this declaration becomes invalid.

Device Name: Lab Centrifuges

Device Model Number(s): C3100 (-E) C3200 (-E)
*Includes all color variations and private label products

Relevant EC Directives: Low Voltage Directive 2014/35/EU
EMC Directive 2014/30/EU
RoHS2 2011/65/EU
RoHS2 2015/863/EU

Harmonized Standards: EN61326-1:2013
EN61010-1:2010

Manufacturer's Name: Benchmark Scientific, Inc
PO Box 709
Edison, NJ 08818

Date: October 1, 2021



Marketing and Quality Control Manager