

Benchmark Scientific Inc. PO Box 709 Edison, NJ 08818

EU DECLARATION OF CONFORMITY

Benchmark Scientific, Inc. hereby declares that the device(s) described below are in conformity with the EC Directives listed here-in as of the date of this declaration. In the event of unauthorized modification of any of the devices listed below, this declaration becomes invalid.

Device Name: Lab Centrifuges

Device Model Number(s): C3100 (-E) C3200 (-E)

*Includes all color variations and private label products

Relevant EC Directives: Low Voltage Directive 2014/35/EU

EMC Directive 2014/30/EU

RoHS2 2011/65/EU

Harmonized Standards: EN61326-1:2013

EN61010-1:2010

Manufacturer's Name: Benchmark Scientific, Inc

PO Box 709 Edison, NJ 08818

Date: January 02, 2019

Marketing Manager

Phone: (908) 769-5555 Fax: (732) 313-7007 Email: info@BenchmarkScientific.com