

Benchmark Scientific Inc.
PO Box 709
Edison, NJ 08818

DECLARATION OF CONFORMITY

Benchmark Scientific, Inc. hereby declares that the device(s) described below are in conformity with the EC Directives listed here-in as of the date of this declaration. In the event of unauthorized modification of any of the devices listed below, this declaration becomes invalid.

Device Name: Accuris™ E3000 UV Transilluminator

Device Model Number(s): E3000, E3000-E and Private label versions of these devices.

Relevant EC Directives: Directive 2006-95-EC
EMC Directive 2004/108/EC (EN 61326-1)

Applicable Standards:

EN ISO 61010-1 (Third Edition):2010
EN ISO 61010-2-030:2010
CISPR 11:2009 + A1:2010
EN 61326-1: 2006, Table 1 (Basic)
EN 61326-2-1: 2006
EN 61000-3-2: 2006 + A1: 2009 + A2: 2009
EN 61000-3-3: 2013
(IEC 61000-4-2: 2008; IEC 61000-4-3: 2006; A1: 2007; A2: 2010;
IEC 61000-4-4:L 2012; IEC 61000-4-5: 2005; IEC 61000-4-6:
2008; IEC 61000-4-11: 2004)

Manufacturer's Name: Benchmark Scientific, Inc.
PO Box 709
Edison, NJ 08818

Date: December 16th, 2017



President