

Benchmark Scientific Inc. PO Box 709 Edison, NJ 08818

## EU DECLARATION OF CONFORMITY

Benchmark Scientific, Inc. hereby declares that the device(s) described below are in conformity with the EC Directives listed here-in as of the date of this declaration. In the event of unauthorized modification of any of the devices listed below, this declaration becomes invalid.

**Device Name:** Thermal Cyclers

**Device Model Number(s):** T5000-96 (-E), T5005-3205 (-E)

\*Including private label versions of these devices.

**Relevant EC Directives:** Low Voltage Directive 2014/35/EU

EMC Directive 2014/30/EU

RoHS2 2011/65/EU RoHS2 2015/863/EU

**Harmonized Standards:** EN61326-1:2013

> EN 61010-2-010:2014 EN61010-1:2010

Manufacturer's Name: Benchmark Scientific, Inc.

PO Box 709

Edison, NJ 08818

Date: January 1st, 2019

Vice President

Email: info@BenchmarkScientific.com Phone: (908) 769-5555 Fax: (732) 313-7007