

Benchmark Scientific Inc. PO Box 709 Edison, NJ 08818

DECLARATION OF CONFORMITY

Benchmark Scientific, Inc. hereby declares that the device(s) described below are in conformity with the EC Directives listed here-in as of the date of this declaration. In the event of unauthorized modification of any of the devices listed below, this declaration becomes invalid.

Device Name: BeadBugTM 3 and BeadBugTM 6 and BeadBlasterTM 24 homogenizers

Device Model Number(s): D1030-E and Private label versions of these devices.

D1036-E and Private label versions of these devices D2400-E and Private label versions of these devices.

Relevant EC Directives: Low Voltage Directive 2006/95/EC

EMC Directive 2004/108/EC Annex II

Harmonized Standards: EN 61010-1: 2010

EN 61010-2-051:2003 EN 61326-1:2013

Manufacturer's Name: Benchmark Scientific, Inc.

PO Box 709 Edison, NJ 08818

Date: Sept 9, 2019

Marketing Manager

Phone: (908) 769-5555 Fax: (732) 313-7007 Email: info@BenchmarkScientific.com