

Benchmark Scientific Inc.  
PO Box 709  
Edison, NJ 08818

## DECLARATION OF CONFORMITY

Benchmark Scientific, Inc. hereby declares that the device(s) described below are in conformity with the EC Directives listed here-in as of the date of this declaration. In the event of unauthorized modification of any of the devices listed below, this declaration becomes invalid.

**Device Name:** Accuris MyGel InstaView Electrophoresis system

**Device Model Number(s):** E1200, E1201 and private label versions thereof

**Relevant EC Directives:** Low Voltage Directive 2006/95/EC  
EMC Directive 2004/108/EC

**Harmonized Standards:** EN 61010-1:2010  
EN 61326-1:2013

**Manufacturer's Name:** Benchmark Scientific, Inc.  
PO Box 709  
Edison, NJ 08818

**Date:** May 5, 2019



Marketing Manager