



# DECLARATION OF CONFORMITY

The undersigned ELMI SIA company declares, that the following instrument(s) conforms the bellow mentioned directives and standards:

Manufacturer:	<b>ELMI SIA</b> Bukultu street 7b, Riga LV-1005, Latvia Phones: (+371) 67558 743, (+371) 67551 934
Equipment name:	RM-2S, RM-2M, RM-2L, V-3
Equipment type:	Laboratory mixer
Directives:	The EMC Directive 2014/30/EU The low voltage directive 2014/35/EU RoHS3 2015/863/EU Directive on the restriction of the use of certain hazardous substances in electrical and electronic equipment WEE 2012/19/EU Directive on waste electrical and electronic equipment In vitro diagnostic medical devices directive 98/79/EC
Applied standards:	<u>LVS EN 61326-1: 2013</u> Electrical equipment for measurement, control and laboratory use EMC requirements. General requirements. <u>LVS EN 61010-1: 2011</u> Safety requirements for electrical equipment for measurement, control and laboratory use. General requirements. <u>LVS EN 61010-2-051: 2015</u> Particular requirements for laboratory equipment for mixing and stirring.

Vitalijs Mironovs  
Technical Director

Maksims Sviridovs  
Manufacturing Director

30.04.2019  
Date

30.04.2019  
Date

