



CE CONFORMITY DECLARATION

The undersigned ELMI SIA company declares, that the following instrument(s) conforms the bellow mentioned directives and standards:

Manufacturer:	ELMI SIA Aizkraukles str. 21-136, Riga LV-1006, Latvia Phone: (+371) 67558 743 Fax: (+371) 67551 934
Equipment name:	RM-2S, RM-2M, RM-2L, V-3
Equipment type:	Laboratory mixer
Directives:	EMC Directive 2004/108/EC Low Voltage Directive 2006/95/EC Medical devices directive 93/42/EEC In vitro diagnostic medical devices directive 98/79/EC
Applied standards:	<u>EN 61326-1: 2013</u> Electrical equipment for measurement, control and laboratory use EMC requirements. General requirements. <u>EN 61010-1: 2010</u> Safety requirements for electrical equipment for measurement, control and laboratory use. General requirements <u>EN 61010-2-051:2015</u> Particular requirements for laboratory equipment for mixing and stirring.

Aleksandrs Sorokins
Commercial Director

Maksims Sviridovs
Technical Director