

Corning Incorporated
Life Sciences

Registered ISO 9001

Product description

Catalog Number TL40-04
PDS reference PDS_Cbor_934 Version _ 1
Product description Corning® Gosselin™ Round Base Tube, 2 mL, PP, Red Straight Plug, Sterile, 1250/Bag, 5000/Case

Sanitary Assurance Level

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10⁻³

Expiration Statement

This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.

Primary body composition

PS

Secondary body composition

N/a

Closure

Red straight plug

Special feature

N/a

Packaging

1250/Bag, 5000/Case

Regulation Statement

We hereby confirm that we do not intentionally use the following substances in the manufacturing of the product :

- LATEX
- BISPHENOL A (CAS No 80-05-7)
- PHTHALATES as per Regulation 1907/2006/EC (SVHC) .


We perform a continuous monitoring of starting material providers regarding the following aspects :

We hereby confirm that we do not intentionally use the following substances in the manufacturing of the product :

- LATEX
- BISPHENOL A (CAS No 80-05-7)
- PHTHALATES as per Regulation 1907/2006/EC (SVHC) .

2) Compliance with:

- EU Food Contact 1935/2010/EC and 10/2011/EC.
- BSE/TSE Free. Product complies with the latest revision of EMA/410/01 "Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via hu
- Compliance to REACH SVHC candidate list (Regulation 1907/2006/EC). They certify that the here above mentioned products do not contain more than 0,1% (w/w) of any substan
- Heavy Metals Free as per directive 94/62/EC; The chemical composition of this product meets the relevant requirements of Directive 92/62/EC (as amended) on packaging and pa



Cedric Maes
Plant Quality Manager, Borre & Chorges
maesc@orning.com