



661 Route 23 South, Wayne, New Jersey 07470  
Tel: 973-694-0500 \* Fax: 973-694-7199  
Internet Address: <http://www.bel-art.com>

## CERTIFICATE OF ANALYSIS

This is to verify that all items noted below are in conformance with the Purchase Order Contract, drawings, specifications, and other applicable documentation; that all process certifications, material certifications, are on file at this or another Bel-Art facility and are available for review by the customer and or Government Representatives upon request.

**Product #:** 36910-0000    **Mfg. Date:** NA    **Lot #:** General

**Description:** "Sterileware" Sterile Transfer Scoop and Bag 2oz.,

**Material:** High-Impact Polystyrene, Poly bag, sterilized in individual polyethylene bags. Scoop Material meets FDA requirements for regulation 21 part 177.1640. White polystyrene meets FDA regulations 21 CFR 177.1640, 178.3297, 175.300 Sample Bag Material meet FDA regulation 21 CFR 177.1520

**Dimensions:** Uses Scoop H36902-0000, 2oz, 17.8cm long, 25mm x 76mm x 38mm (bowl), 4.5" x 9" Resealable Poly Bag 2.5 mil thick, Usable depth approx. 5", Approx. volume 0.125 gal.

**Other:** Irradiation measured absorbed dose (avg.)= 23.5 to 50.0 kGy , No testing is done for Pyrogens / End toxins at this time. Based on the process and materials involved in the manufacturing of this product, this product is considered to be BSE/TSE protein free. AAMI guidelines are followed. Product should remain in its irradiated state as long as outer packaging remains completely intact. Product should be stored in room temperature, dry location. Based on the information provided to us this product does not contain or come in contact with lead, latex, or other contaminants. To the best of our knowledge based on information provided to us this product does not contain or come in contact with any animal origin.  
Manufactured: Bel-Art Products Pocomoke City MD, USA

Reviewed and Approved By:

**Signature:** 

**Printed Name:** Carl Sisco

**Title:** Technical Support Manager

**Date:** 6/28/2018