

CERTIFICATE OF ANALYSIS

This is to verify that all items noted below are in conformance with the Purchase Order Contract, drawings, specifications, and other applicable documentation; that all process certifications, material certifications, are on file at this or another Bel-Art facility and are available for review by the customer and or Government Representatives upon request.

Product #: 36933-0001 **Mfg. Date:** NA **Lot #:** General

Description: Sterileware® Economy Sampling Scoops

Material: High-Impact Polystyrene Natural, sterilized in individual polyethylene bags.
Materials meet FDA requirements for regulation 21 part 177.1640
White polystyrene Colorant FDA regulations 21 CFR 177.1640, 178.3297, 175.300

Dimensions: 5.06" L x 1.73"W x 1.02" H, approx. bowl at top 3.43" L x 1.55" W x
96" D, approx. bowl bottom 2.7" L x .86" W, Vol. 63ml

Other: Gamma Treatment: 23.5-50.0 kGy
No testing is done for Pyrogens / End toxins at this time.
Based on the process and materials involved in the manufacturing of this product
this product is considered to be BSE/TSE protein free. AAMI guidelines are followed
Based on the information provided to us this product does not contain or come in
contact with lead, latex, or other contaminants.

Reviewed and Approved By:

Signature: 

Printed Name: Carl Sisco

Title: Technical Support Manager

Date: 7/2/2018