



661 Route 23 South, Wayne, New Jersey 07470
Tel: 973-694-0500 * Fax: 973-694-7199
Internet Address: <http://www.bel-art.com>

This is to verify that all items noted below are in conformance with the Purchase Order Contract, drawings, specifications, and other applicable documentation; that all process certifications, material certifications, are on file at this or another Bel-Art facility and are available for review by the customer and or Government Representatives upon request.

Product #: 36902-0006 **Mfg. Date:** na **Lot #:** General

Description: "Sterileware® Sense-able Scoops® Sterile Sampler Scoops (BLUE)
Material: High-Impact Polystyrene, Blue PP Colorant, EEA-SS Grey Resin, FDA Grade
Resin meets FDA regulation 21 CFR 177.1520
Polyethylene bag materials meet FDA requirements for regulation 21 part 177.1640

Dimensions: 2oz., OD 76mm L x 41mm W x 29mm Maximum D (bowl)
Overall 203mm long, Handle 127mm, 9.6mm W x 9.0mm H, tolerance +/- 0.020"

Other: Gamma Irradiation exposure dose = min. 23.5 and max. 50.0 kGy
ANSI/AAMI/ISO 11137:2006 method 1 guideline are followed.
Metal and X-ray detectable. To the best of our knowledge based on information provided to us this product does not contain or come in contact with any animal origin. Based on the process and materials involved in the manufacturing of this product, this product is considered to be BSE/TSE protein free.
Product should remain in its irradiated state as long as outer packaging remains completely intact. Product should be stored in room temperature, dry location. Based on the information provided to us this product does not contain or come in contact with lead, latex, Mercury, Melamine, formaldehyde or other contaminants.
Manufactured: Bel-Art Products, Pocomoke City, MD USA

Reviewed and Approved By:

Signature: *Lil Elligson*

Printed Name: Lil Elligson

Title: Senior Quality Associate

Date: 10/17/2017

SUBSIDIARIES: MADDAK, INC * APPLIED COATING, INC
ADDITIONAL PLANTS IN POCOMOKE, MD * KANSAS CITY, MISSOURI