Declaration of conformity

INTEGRA Biosciences AG – 7205 Zizers, Switzerland

declares on its own responsibility that the devices

Description	Models
PIPETBOY acu 2	**155000, 155015, 155016, 155017, 155018, 155019**
PIPETGIRL	**155021**

comply with:

EU Directives (DoW: Date of Withdrawal)	Before DoW	DoW	After DoW
Low Voltage Equipment	**2006/95/EC**	20.04.2016	**2014/30/EU**
Electromagnetic Compatibility	**2004/108/EC**	20.04.2016	**2014/35/EU**
Restriction of Hazardous Substances	**2011/65/EU**		
Waste Electrical and Electronic Equipment	**2012/19/EU**		
Battery Directive	**2006/66/EC**		

EU Regulations

Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH)	**1907/2006**
Capacity Labelling of Portable Secondary Batteries	**1103/2010**
Ecodesign - Power supplies	**278/2009**

Standards for EU

Safety requirements for electrical equipment for measurement, control and laboratory use - General requirements.	**EN 61010-1: 2010**
Electrical equipment for measurement, control and laboratory use - EMC requirements.	**EN 61326-1: 2006**

Standards for Canada and USA

Safety requirements for electrical equipment for measurement, control and laboratory use - General requirements.	**CAN/CSA-C22.2 No. 61010-1**
Safety requirements for electrical equipment for measurement, control and laboratory use - General requirements.	**UL 61010-1**
Operation is subject to the following two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.	**Part 15 of the FCC Rules**

Zizers, January 6, 2016

Elmar Morscher
CEO

Thomas Neher
Quality Manager